



# Declaration of Conformity

MANUFACTURER



: Bioneer Corporation  
8-11, Munpyeongseo-ro, Daedeok-gu,  
Daejeon, 34302  
Republic of Korea

EUROPEAN  
REPRESENTATIVE



: MT Promedt Consulting GmbH  
Altenhofstr. 80  
D-66386 St. Ingbert, Germany

PRODUCT

: *ExiPrep*<sup>TM</sup>48 Dx,  
Fully Automated Nucleic Acid Extraction System

CATALOG NO.

**REF** A-5150

EDMA Code/Term  
CLASSIFICATION

: 21 01 41 Small Automated CC Analyzer  
: Others  
(Neither Listed in Annex II of IVDD, Nor self-testing device)

CONFORMITY  
ASSESSMENT ROUTE

: IVDD ANNEX III

*We herewith declare that the above mentioned products meet the provisions of the council directive 98/79/EC for in vitro diagnostic medical device. All supporting documentation is retained under the premises of the manufacturer.*

STANDARDS APPLIED

: EN 61010-1:2010, EN 61010-2-081:2015,  
EN 61010-2-010:2014, EN61010-2-101:2002,  
EN61326-1:2013, EN61326-2-6:2013  
EN 61000-4-2: 2009, EN 61000-4-3:2006/A1:2008/A2:2010,  
EN 61000-4-4:2012, EN 61000-4-5:2014,  
EN 61000-4-6:2009, EN 61000-4-11:2004  
EN ISO 14971:2012, EN ISO 13485:2016, EN 62304:2006,  
EN 62366:2008, EN ISO15223-1:2016, EN ISO 18113-1:2011,  
EN ISO 18113-3:2011

START DATE OF CE MARKING: 2018-10-05

PLACE, DATE OF ISSUE : Daejeon, Republic of Korea

SIGNATURE:

Han-Oh PARK, Ph.D.

DATE: 2018-10-05