

Declaration of Conformity



Description of Device: **AUTO REFRACTO-KERATOMETER**

Model Name: **URK-800F**

Classification: **Class IIa**

**according to the Rule 10 of Classification Criteria,
annex IX, MDD 93/42/EEC amended by 2007/47/EC**

**Conformity Assessment Route : Annex II (excluding section4), MDD 93/42/EEC As
amended by 2007/47/EC**

Applied Standards: **IEC 60601-1: 2005+CORR.1(2006) + CPR.2(2007)
IEC 60601-1-2:2007(Third Edition)**

Manufacturers Name: **UNICOS Co., Ltd.**

49, Gwarye-ro 22beon-gil, Jung-Gu, Deajeon, Korea

TEL:+82-42-581-0008 FAX:+82-42-825-8041

Website:<http://www.e-unicos.com>

European representative : Ellegi Medical Optics s.r.l. Via Pisciarelli 79- 80078, Pozzuoli(NA)

TEL:081.5499645 - 081.5498044 Partita I.V.A. 05354730631

We hereby declare that the complies with the Medical Devices Directive 93/42/EEC As amended by 2007/47/EC (and its relevant transposition into the national laws of the Member States in which the device is intended to be placed on the market) using Annex II as the conformity assessment procedure via SGS(NB 0120) as the Notified Body.

Place, Date of issue : Jung-Gu, Daejeon, Korea Jan. 24, 2017



UNICOS CO., LTD

K.C. KIM

K.C KIM PRESIDENT

Ki-chang, Kim/ President

On behalf of UNICOS.Co.,Ltd