

***Certificate of Analysis***

Product Name: Haptoglobin Kit  
OEM Part Number: 09D911T21  
OEM Lot Number: 15846UN21  
Storage: 2 - 8°C  
Qty Manufactured: 145 Kits

Abbott List Number: 09D9121  
Abbott Commodity Number: 39-6827  
Expiration Date: 2022-10-06  
Package Insert No./Rev. 306773 / R04

Specification	Specification Limit	Result
Barcode Scan	Acceptable verification on c8000	Complies Pass
Bio-Rad Control L1	65 - 79 mg/dL	75 Pass
Bio-Rad Control L1	Lot Number	68951
Bio-Rad Control L2	139 - 170 mg/dL	159 Pass
Bio-Rad Control L2	Lot Number	68952
Bio-Rad Control L3	209 - 256 mg/dL	242 Pass
Bio-Rad Control L3	Lot Number	68953
Control Level 4: % Difference	-10 - 10 %	2 Pass
Micro: Beg Sample (R1)	<30 CFU/mL	0 Pass
Micro: End Sample (R1)	<30 CFU/mL	0 Pass
Micro: Beg Sample (R2)	<30 CFU/mL	0 Pass
Micro: End Sample (R2)	<30 CFU/mL	0 Pass

For in-vitro diagnostic use.

This lot meets all current release specifications per quality document MS.MASTER342.

The above product has been manufactured in compliance with Quality Systems Regulations, passed all Thermo Fisher final release specifications, and has met all Abbott acceptance requirements.

This product contains R1 and R2 components. The component name, lot number, and expiration date matches the kit name, lot number, and expiration date listed on this COA.

This product contains animal source material. The caprine material used in the manufacture of this product is derived from goat antiserum collected from U.S. sourced goats.



Quality Representative

Date: 2021-10-11