

# **Hướng dẫn sử dụng bản gốc**

Tài liệu được cung cấp bằng tiếng Anh

Tài liệu được xác nhận bằng chữ ký số

Tp Hà Nội ngày 18 tháng 04 năm 2022

REF C88007M



2 × 50 Tests



### INTENDED USE

The iFlash-EB VCA IgM assay is a paramagnetic particle chemiluminescent immunoassay (CLIA) for the qualitative determination of antibody to viral capsid antigen of Epstein-Barr virus in human serum and plasma using the iFlash Immunoassay Analyzer.

### SUMMARY AND EXPLANATION

Epstein-Barrvirus (EBV) is generally acquired by oral transmission of infected saliva and genital secretions. It is a worldwide infectious agent that is the classic cause of infectious mononucleosis (IM). IM is often diagnosed based on the presentation of characteristic symptoms of sore throat, lymphadenopathy and fever and intimately associated with several human tumors, including African Burkitt lymphoma, nasopharyngeal carcinoma, Hodgkin's disease, and B-cell lymphomas.

EBV infects B cells of immune system and epithelial cells. Most people become infected with EBV and gain adaptive immunity. The diagnosis of EBV infections can be characterized based on several distinct EBV-associated antigen systems. The acute phase of infectious mononucleosis is characterized by rapid IgM and IgG antibody responses to VCA.

### ASSAY PRINCIPLE

The iFlash-EB VCA IgM assay is an indirect immunoassay.

- 1<sup>st</sup> incubation: EB VCA IgM in the sample and EB VCA antigen-coated paramagnetic microparticles react to form a complex.
- Wash: The unbound materials are washed away from the solid phase in a magnetic field.
- 2<sup>nd</sup> incubation: Acridinium-ester-labeled anti-human IgM conjugate is added to form a reaction mixture.
- Another Wash.
- Trigger of signal: The Pre-Trigger and Trigger Solutions are added to the reaction mixture. The resulting chemiluminescent reaction is measured as relative light units (RLUs).
- A direct relationship exists between the amount of EB VCA IgM in the sample and the RLUs detected by the iFlash optical system.
- Results are determined by comparing the RLUs of each sample to the cutoff signal obtained from a previous calibration.

### REAGENTS

Reagent kit, 100 tests, 2 packs, 50 tests/pack

R1	EB VCA antigen-coated paramagnetic microparticles, 3.5 mL/pack, 0.05% ProClin 300.
R2	Acridinium-ester-labeled anti-human IgM conjugate, 6.5 mL/pack; 0.05% ProClin 300.

R3	Sample diluent, phosphate buffer, 6.5 mL/pack, 0.05% ProClin 300.
CAL1	Calibrator 1, 1 bottle, 1.0 mL, Tris buffer with protein stabilizers, 0.05% ProClin 300.
CAL2	Calibrator 2, 1 bottle, 1.0 mL, EB VCA IgM in Tris buffer with protein stabilizers, 0.05% ProClin 300.

### MATERIALS REQUIRED (BUT NOT PROVIDED)

REF C89999/C89959/C89949, iFlash Pre-Trigger Solution: hydrogen peroxide solution.

REF C89998/ C89958/ C89948, iFlash Trigger Solution: sodium hydroxide solution.

REF C89997, iFlash Wash Buffer: phosphate buffered saline solution with 0.05% ProClin 300.

REF C80001, iFlash Wash Buffer (10×): phosphate buffered saline solution with 0.05% ProClin 300.

REF C89996, reaction vessels.

### WARNINGS AND PRECAUTIONS

IVD For in vitro diagnostic use

- The calibrator (CAL2) has been prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg, anti-HCV, and anti-HIV-1/2 by approved methods.
- No known test method can offer the complete assurance that products derived from human sources will not transmit infection. Therefore, all humanized materials should be considered potentially infectious.
- Exercise the normal precautions required for handling all laboratory reagents.
- Disposal of all waste material should be in accordance with local guidelines.
- Wear gloves when handling specimens or reagents.
- Clean and disinfect all spills of specimens or reagents using a suitable disinfectant.
- iFlash Trigger solution contains sodium hydroxide (NaOH) and should be avoided contacting eyes.
- For further information on warnings and precautions, see Annex B.

### REAGENT HANDLING

- The reagents may not be used after the stated expiration date.
- Avoid the formation of foam with all reagents.
- The reagents in the pack and calibrators are ready for use.
- Close the bottles of calibrator right after calibration and store at 2-8°C.
- Do not pool reagents within a reagent kit or between reagent kits.
- Prior to loading the iFlash-EB VCA IgM reagent pack on the system for the first time, resuspend the microparticles by inverting the reagent pack slightly.
- For further information on reagent handling precautions during system operation, refer to the iFlash system operating instruction.

## STORAGE AND STABILITY

### Storage:

- Store at 2-8°C in an upright position.
- The kit may be used immediately after removal from 2-8°C storage.

### Stability:

- Unopened at 2-8°C: up to the stated expiration date.
- Opened at 2-8°C: 28 days.
- Store on-board: 28 days.

## SPECIMEN COLLECTION AND PREPARATION

- Serum or plasma (lithium heparin, sodium heparin potassium EDTA, and sodium citrate) are the recommended samples. Other anticoagulants have not been validated for use with the iFlash-EB VCA IgM assay.
- Ensure that complete clot formation in serum specimens has taken place prior to centrifugation.
- Centrifuge the specimens.
- Store specimens at room temperature (20 to 25°C) for no longer than 8 hours.
- If the testing will not be completed within 8 hours, refrigerate the samples at 2 to 8°C.
- If the testing will not be completed within 14 days, or for shipment of samples, freeze at -20°C or colder.
- Frozen specimens must be mixed thoroughly after thawing.
- The samples may be frozen for 3 times in maximum.
- Centrifuge specimens with a lipid layer on the top, and transfer only the clarified specimen without the lipemic material.
- Ensure that residual fibrin and cellular matter have been removed prior to analysis.
- Use with caution when handling patient specimens so as to prevent cross-contamination.
- Do not use heat-inactivated samples.
- Ensure that the patient samples, calibrators and controls are at ambient temperature (20-25°C) before measurement.
- Due to the possible evaporation, specimens and calibrators on the analyzers should be measured within 2 hours.

## ASSAY PROCEDURE

- Refer to the system operating instruction or the online help system for detailed information on preparing the system.
- The test-specific parameters stored in barcode on the reagent pack are read in. In case the barcode cannot be read, enter the sequence numbers.
- Carry out calibration, if necessary.
- Place the calibrators CAL1 and CAL2 in the calibrator rack in the sample zone. Only keep calibrators open during calibration.

- Test application.
- Load samples (20 µL of sample is needed for each determination in addition to the sample container and system dead volumes).
- Click RUN, the iFlash System performs all the functions automatically and calculates the results.

## CALIBRATION

- Every iFlash-EB VCA IgM reagent kit has a QR code label containing the specific information for calibration of the particular reagent lot.
- To perform an iFlash-EB VCA IgM calibration, test CAL1 and CAL2 in duplicate, and the predefined master curve is adapted to the analyzer.
- Once an iFlash-EB VCA IgM calibration is accepted and stored, all subsequent samples may be tested without further calibration unless:
  - After 28 days when using the same reagent lot.
  - A new lot reagent kit is used.
  - Controls are out of range.
  - Required by pertinent regulations.

## QUALITY CONTROL

Quality control materials should be run as single determinations at least once every 24 hours when the test is in use, once per reagent kit and after every calibration. Include commercially available quality control materials that cover at least two levels of analyte. Follow manufacturer's instructions for reconstitution and storage. Each laboratory should establish mean values and acceptable ranges to assure proper performance. Quality control results that do not fall within acceptable ranges may indicate invalid test results.

## RESULT

### Calculation:

The iFlash system automatically calculates the analyte concentration of each sample. The results are given in U/mL

### Interpretation of Results

- Nonreactive: < 20.0 U/mL
- Indeterminate: ≥20.0 U/mL - < 40.0 U/mL
- Reactive: ≥40.0 U/mL

## LIMITATIONS

- The iFlash-EB VCA IgM assay is limited to the determination of EB VCA IgM in human serum or plasma (lithium heparin, sodium heparin, potassium EDTA, and sodium citrate). It has not been validated for use with other types of plasma.
- The use of serum separator (gel) blood collection tubes has been validated for use with this assay. However, it is not possible to survey all manufacturers or tube types.
- If the results are inconsistent with clinical evidence, additional testing is suggested to confirm the result.

- For diagnostic purposes, the results should be interpreted in light of the total clinical presentation of the patient, including symptoms, clinical history results.
- Specimens from heparinized patients may be partially coagulated and erroneous results could occur due to the presence of fibrin.
- The results from an alternative assays (i.e. EIA or RIA) may not be equivalent and cannot be used interchangeably.
- The assay is unaffected by icterus (bilirubin < 30 mg/dL), hemolysis (Hb < 1,500 mg/dL), lipemia (Intralipid < 1,500 mg/dL) and total serum protein (< 10 g/dL).
- No interference was observed from rheumatoid factors up to a concentration of 2,000 IU/mL.
- No interference was observed from anti-nuclear antibodies up to a concentration of 500 AU/mL.
- No interference was observed from HAMA up to a concentration of 600 ng/mL.

### PERFORMANCE CHARACTERISTICS

Below are the representative performance data, and the results obtained in individual laboratories may differ

#### Precision

The iFlash-EB VCA IgM is designed to have a precision of  $\leq 10\%$  total CV.

The precision of iFlash-EB VCA IgM was determined using EB VCA IgM reagents, samples and controls. Two serum samples, consisting low and median concentration of EB VCA IgM were assayed.

The within-run precision was determined by testing each sample in replicates of 10 (n = 10), and calculating percent coefficient of variation (%CV). The results of the study are shown below:

Sample	Mean (U/mL)	SD	%CV
1	54.6	1.80	3.30%
2	104.8	5.52	5.27%

The between-run precision was determined by testing each sample in duplicate, two separate runs daily for 20 days (n = 80), and calculating percent coefficient of variation (%CV). The results of the study are shown below:

Sample	Mean (U/mL)	SD	%CV
1	53.1	4.13	7.78%
2	102.9	5.10	4.96%

#### Analytical Specificity

The analytical specificity of iFlash-EB VCA IgM assay was evaluated with viral antibody specimens. The nonreactive EB VCA IgM status of each specimen was verified using a commercially available EB VCA IgM assay.

Clinical Category	Number of Specimens	iFlash EB VCA IgM Nonreactive
EB VCA IgG	3	3
EB VCA IgA	9	9
VZV IgM	6	6

Toxo IgM	4	4
Rubella IgM	5	5
CMV IgM	12	12
HSV-1/2 IgM	18	18
Total Samples	57	57

#### Relative Specificity

Relative sensitivity of iFlash-EB VCA IgM assay was determined by testing samples that were found reactive in a commercially available EB VCA IgM assay. A total of 240 samples including sequential and single samples were tested with iFlash-EB VCA IgM assay.

Clinical Category	Number of Specimens	Reactive Specimens	Relative Sensitivity
EB VCA infected	240	237	98.75%

#### Relative Sensitivity

Relative specificity of iFlash-EB VCA IgM assay was determined by testing samples that were found nonreactive in a commercially available EB VCA IgM assay. A total of 744 samples including blood donors and hospitalized patients were tested with iFlash-EB VCA IgM assay.

Clinical Category	Number of Specimens	Non-reactive Specimens	Relative Specificity
Blood donors	744	742	99.73%
Hospitalized patients			

#### REFERENCES

1. deThe G, Ho JHC, Muir CS. Nasopharyngeal carcinoma. In: Evans AS, ed. Viral Infection of Human: Epidemiology and Control. New York: John Willey & Sons, 1982; 621-652.
2. Lung ML, Chan KH, Lam WP et al. In-situ detection of Epstein-Barr virus markers in nasopharyngeal carcinoma tissue. *Oncology* 1989; 46: 310-317.
3. de Schryver A, Friberg S, Klein Wet al. EBV associated antibody patterns in carcinoma of the post-nasal space. *Clin Exp Immunol*, 1969; 5; 443-459.
4. deThe G. Virology and immunology of nasopharyngeal carcinoma: Present situation and outlook. In: Biggs PM, deThe G, Payne LN, eds. *Oncogenesis and Herpesvirus*. International Agency for Research on Cancer Science Publication No 2. 1972; 275-284.
5. Henle W, Henle G, Ho JHC et al. Antibodies to Epstein-Barr virus in nasopharyngeal carcinoma, other head and neck neoplasms and control groups. *J Natl Cancer Inst* 1970; 44: 225-231.

SHENZHEN YHLO BIOTECH CO., LTD.



Building 1, YHLO Biopark, Baolong 2nd Road, Baolong Subdistrict, Longgang District, 518116 Shenzhen, PEOPLE'S REPUBLIC OF CHINA



Wellkang Ltd (www.CE-marking.eu)

Suite B, 29 Harley St., London W1G 9QR, UK

(Reaction mass of: 5-chloro-2-methyl-4-isothiazolin [EC no. 247-500-7] and 2-methyl-4-isothiazolin-3-one [EC no. 200-239-6] (3:1))

- Hazard Statement:
  - H317: May cause an allergic skin reaction.
  - H319: Causes serious eye irritation.
  - H410: Very toxic to aquatic life with long-lasting effects.
- Precautionary Statement:
  - P261: Avoid breathing dust/fume/gas/mist/vapours/spray.
  - P264: Wash hands thoroughly after handling.
  - P272: Contaminated work clothing should not be allowed out of the workplace.
  - P280: Wear protective gloves/protective clothing/eye protection/face protection.
  - P305+P351+P338: IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.
  - P337+P313: If eye irritation persists: Get medical advice/attention.
  - P333+P313: If skin irritation or rash occurs: Get medical advice/attention.
  - P302+P352: IF ON SKIN: Wash with plenty of soap and water.
  - P321: Seek immediate care from a doctor.
  - P363: Wash contaminated clothing before reuse.
  - P273: Avoid release to the environment.
  - P391: Collect spillage.
  - P501: Dispose of contents/container in a safe way.

## ANNEX A:

### Explanation of abbreviation

Abbreviation	Explanation
	Product No.
	Calibrator
	Reagent
	Number of tests
	Manufactured by
	EU Representative
	EC Declaration of Conformity
	Caution
	Instructions for use
	In vitro diagnostic medical device
	Lot No.
	Date of manufacture
	Expiry date
	Biohazard Symbol
	Pictograms for Caution
	Pictograms for Hazardous to the aquatic environment

## ANNEX B:

### WARNINGS AND PRECAUTIONS (Proclin 300)

- Hazardous Component: 0.05% Proclin 300