



# Certificate of Analysis

Abbott Ireland Diagnostics Division,  
Lisnamuck,  
Longford,  
Co. Longford,  
Ireland

**Product Name:** Activated Alanine Aminotransferase      SCP / revision      8L920024/003  
**List number:** 08L9224  
**Lot number:** 38437UD00  
**Reagent Lot Number:** 38437UD00  
**Expiry Date:** 2023-01-28

Component name	List Number	Lot Number	Production date	Expiry date
A-ALTR1	08L92R	38438UD00	2021-12-28	2023-01-28
A-ALTR2	08L92T	38439UD00	2021-12-28	2023-01-28

## Acceptance Criteria: Grand Mean Concentration (U/L)

Description	Lower Limit	Upper Limit	Value	Result
REF Control 1	22.00	28.02	24.92	Acceptable
REF Control 2	74.75	88.03	80.55	Acceptable
REF Control 3	195.40	230.12	210.20	Acceptable
REF GC3 L5	760.19	929.13	848.75	Acceptable

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality:

CARMIL FABIAN (Fabian)  
Printed name and signature

Date:

31 March 22