Zimmer MedizinSysteme GmbH
Junkersstraße 9
89231 Neu-Ulm / Germany

Tel.: +49 731 / 9761-0
Fax: +49 731 / 9761-118

info@zimmer.de
www.zimmer.de

Declaration of Conformity

for the electrostimulation, ultrasound therapy and suction wave device family **Soleoline**

We declare under our sole responsibility, that the product line

Soleoline
consisting of the devices

Soleo **Sono**
Soleo **SonoStim**
Soleo **Galva**
VacoS

GIÁM ĐỐC
Triệu Đức Thành

bears the CE mark

CE 0123

according to the Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4) and fulfils the essential requirements of annex I of this directive.

The device belongs to class IIa according to annex IX rule 9 of this directive.

The conformity of the quality management system is certified and controlled by TÜV SÜD Product Service GmbH, Ridlerstraße 65, 80339 Munich, Notified Body Code: 0123.

This declaration is valid from 2020-08-06 until 2024-05-26.

Stefan Leinweber
Quality Management
Medical Device Safety Officer

Neu-Ulm, 2020-08-06