

## Declaration of Conformity

for the shortwave diathermy device type **ThermoPro**

We declare under our sole responsibility, that the device

**ThermoPro**

bears the CE mark

**CE** 0123



**GIÁM ĐỐC**

*Triệu Đức Thành*

according to the Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4) and fulfils the essential requirements of annex I of this directive.

The device belongs to class IIa according to annex IX rule 9 of this directive.

The conformity of the quality management system is certified and controlled by TÜV SÜD Product Service GmbH, Ridlerstraße 65, 80339 Munich, Notified Body Code: 0123.

This declaration is valid from 2020-08-07 until 2024-05-26.

A handwritten signature in black ink, appearing to read 'Stefan Leinweber'.

Stefan Leinweber  
Quality Management  
Medical Device Safety Officer

Neu-Ulm, 2020-08-07