



# Certificate of Analysis

Abbott Ireland Diagnostics Division,  
Lisnamuck,  
Longford,  
Co. Longford,  
Ireland

**Product Name:** Alinity i Prolactin Controls

List number: 07P6610

Lot number: 36187UD00

Expiry Date: 2023-01-05

SCP / revision

7P660010/002

Component name	List Number	Lot Number	Production date	Expiry date
Alinity i Prolactin Control L	7P66L	36184UD00	2022-01-05	2023-01-05
Alinity i Prolactin Control M	7P66M	36185UD00	2022-01-05	2023-01-05
Alinity i Prolactin Control H	7P66N	36186UD00	2022-01-05	2023-01-05

## Acceptance Table: Mean % Difference (SMP vs. REF ng/mL)

Description	Lower Limit	Upper Limit	Value	Result
Low Control	-4.4	10.3	-1.1	Acceptable
Medium Control	-12.5	9.5	1.2	Acceptable
High Control	-9.5	9.8	-2.4	Acceptable

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality:

MACEJ SOBIEZYK Maj Sam  
Printed name and signature

Date:

31 Jan 22