



Certificate of Analysis

Abbott Ireland Diagnostics Division,
Lisnamuck,
Longford,
Co. Longford,
Ireland

Product Name: Alinity i Urine NGAL Controls

List number: 09P2910

Lot number: 30123UD00

Expiry Date: 2023-01-01

SCP / revision

9P290010/002

Component name	List Number	Lot Number	Production date	Expiry date
Alinity i Urine NGAL Control L	9P29L	30117UD00	2021-07-01	2023-01-01
Alinity i Urine NGAL Control M	9P29M	30118UD00	2021-07-01	2023-01-01
Alinity i Urine NGAL Control H	9P29N	30119UD00	2021-07-01	2023-01-01

Acceptance Table: Mean S/C Reported Result

Description	Lower Limit	Upper Limit	Value	Result
Low Control	0.900	1.100	0.981	Acceptable
Medium Control	0.900	1.100	0.971	Acceptable
High Control	0.900	1.083	0.983	Acceptable

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality:

Kieran Green
Printed name and signature

Date:

20 July 21