



Tel: 508-660-2221  
 Fax: 508-660-2224  
 Email: info@htmed.com  
 Web: http://htidiagnostics.com  
 High Technology, Inc.

Catalog No  
 ELTP-2014  
 ELT5-1014A

**INTENDED USE**

To measure ion concentrations of Chloride in the sample

**MATERIALS PROVIDED**

Electrode, Chloride (Cl)

**PREPARATION**

Ready to use

**STORAGE AND STABILITY**

Store the reagent at room temperature (0 - 40°C)

**WARNINGS AND PRECAUTIONS**

1. For *in vitro* diagnostic use.  
*CAUTION: In vitro* diagnostic reagents may be hazardous. Handle in accordance with good laboratory procedures which dictate avoiding ingestion, and eye or skin contact. The reagent is an acid solution. Flush with water when contact occurs.
2. Specimens should be considered infectious and handled appropriately.
3. Avoid ingestion. Flush with water when contact occurs.










**REAGENT DETERIORATION**

The reagent should be a clear. If turbidity or precipitation has occurred the reagent should be discarded.

**Manufactured By**  
 High Technology, Inc.  
 20 Alice Agnew Dr.  
 North Attleboro, MA 02763 USA  
 Tel: 1-508-660-2221  
 E-mail: info@htmed.com  
 www.htmed.com, http://htidiagnostics.com

EC	REP
----	-----

EMERGO EUROPE  
 Prinsessegracht 20  
 2514 AP The Hague  
 The Netherlands  
 +31 70 345 8570

	Manufactured by		Used by
	Catalog number		In vitro diagnostic
	Temperature limitation		Authorized representative in the European Community
	Lot number		Consult instructions for use
	This product conforms to the directive 98/79/EC (IVD-directive)		