



Tel: 508-660-2221
 Fax: 508-660-2224
 Email: info@htmed.com
 Web: http://htidiagnostics.com
 High Technology, Inc.

ELT5-1010A
 ELTP-2010

INTENDED USE

To measure ion concentrations of Magnesium in the sample

MATERIALS PROVIDED

Electrode, Magnesium (Mg)

PREPARATION

Ready to use

STORAGE AND STABILITY

Store the reagent at room temperature (0 - 40°C)

WARNINGS AND PRECAUTIONS

1. For *in vitro* diagnostic use.
CAUTION: In vitro diagnostic reagents may be hazardous. Handle in accordance with good laboratory procedures which dictate avoiding ingestion, and eye or skin contact. The reagent is an acid solution. Flush with water when contact occurs.
2. Specimens should be considered infectious and handled appropriately.
3. Avoid ingestion. Flush with water when contact occurs.










REAGENT DETERIORATION

The reagent should be a clear. If turbidity or precipitation has occurred the reagent should be discarded.

Manufactured By
 High Technology, Inc.
 20 Alice Agnew Dr.
 North Attleboro, MA 02763 USA
 Tel: 1-508-660-2221
 E-mail: info@htmed.com
 www.htmed.com, http://htidiagnostics.com

EC	REP
----	-----

EMERGO EUROPE
 Prinsessegracht 20
 2514 AP The Hague
 The Netherlands
 +31 70 345 8570

	Manufactured by		Used by
	Catalog number		In vitro diagnostic
	Temperature limitation		Authorized representative in the European Community
	Lot number		Consult instructions for use
	This product conforms to the directive 98/79/EC (IVD-directive)		