

		Ningbo Greetmed Medical Instruments Co., Ltd. Declaration of Conformity				Prepared by	Dong Jingjing
						Checked by	Chen Xiaoxia
						Approved by	Li Guirong
Edition	A/3	Doc. No	GT/CE-96-01	Effective Date	2019-12-17	Total page	Page 1 of 1

Manufacturer:

Ningbo Greetmed Medical Instruments Co., Ltd.
 Address: 16F-1, Building 1, No.98 Chuangyuan Road, Hi-Tech Zone, 315042 Ningbo, Zhejiang Province, PEOPLE'S REPUBLIC OF CHINA
 Tel: 86-574-87739070 Fax: 86-574-87722360

European Representative:

Shanghai International Holding Corp. GmbH (Europe)
 Address: Eiffestrasse 80, 20537 Hamburg, Germany

Product Name: Tracheostomy Tubes

Size: I.D.3.0, I.D.3.5, I.D.4.0, I.D.4.5, I.D.5.0, I.D.6.0, I.D.6.5, I.D.7.0, I.D.7.5, I.D.8.0, I.D.8.5, I.D.9.0, I.D.9.5, I.D.10.0

GMDN Code: 35404

Classification (MDD, Annex IX): II b (Rule 8 of Annex IX)

Conformity Assessment Route: Annex II .3

We herewith declare that the above mentioned products meet the transposition into national law, the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer.

Ningbo Greetmed Medical Instruments Co., Ltd. is exclusively responsible for Declaration of Conformity

DIRECTIVES

General applicable directives:

Medical Device Directive: COUNCIL DIRECTIVE 93/42/EEC of 14 June 1993 concerning medical devices (MDD 93/42/EEC).

Notified Body: TÜV SÜD Product Service GmbH, Ridlerstr 65, 80339 MÜNchen, Germany

Identification number: CE0123

(EC) Certificate(s): G1 073283 0045 Rev. 00

Expire date of the Certificate: 2021-04-17

Place, Date of Issue: Ningbo, 2021-04-01

Signature: 

Name: Li Guirong

Position: General Manager

