

# Xenosys.Co.,Ltd.

## *DECLARATION OF CONFORMITY*

Product submitted ..... : Medical Head Lamp, Class 1 device  
Model: L2S14C  
Rated Input: 100-240 V~; 50-60 Hz; 0.3 A

Applicant(Manufacturer).... : Xenosys.Co.,Ltd  
2 Venture-ro, 100 beon-gil, Yeonsu-gu, Incheon 406-840 Republic of Korea,

Testing Laboratory ..... : SGS Korea Co., Ltd. (Gunpo Laboratory)

Test Report Number(s)..... : EMC : F609501/RF-EMG005168  
Safety : F690501/RF-SAF006821


Order Number(s)..... : G-44-2014-00684

Specification Requested .... : EMC : EN 60601-1-2:2007  
EN55015:2006/A2:2009  
EN55011:2009/A1:2010

Safety : IEC60601-1-1  
IEC60601-1-6

### Conclusion

Based on a review of the test report, this apparatus meets the requirements of the above standards and hence fulfills the requirements of MDR (EU) 2017/745



.....  
Ha-deuk Lee / President of Xenosys Co., Ltd.

.....  
April 6, 2020