

Xenosys.Co.,Ltd.

DECLARATION OF CONFORMITY

Product submitted: Medical Head Lamp, Class 1 Device
Model: L2SWE
Ratings: AC/DC Adaptor: Input: 100-240 V~, 50-60 Hz; 0.3-0.15 A
Output: 5Vd.c., 2.2A
ME equipment: 3.7Vd.c., 350mA

Applicant(Manufacturer)....: Xenosys.Co.,Ltd
2 Venture-ro, 100 beon-gil, Yeonsu-gu, Incheon 22013 Republic of Korea,


Testing Laboratory: SGS Korea Co., Ltd. (Gunpo Laboratory)

Test Report Number(s).....: EMC : F690501-RF-EMG000195
Safety : F690501/RF-SAF000612
Usability: F690501/RF-SAF000613

Specification Requested ...: EMC : IEC 60601-1-2:2014
Safety : IEC 60601-1:2005, COR1:2006, COR2:2007, AMD1:2012
Usability: IEC 60601-1-6:2010, AMD1:2013

Conclusion

Based on a review of the test report, this apparatus meets the requirements of the above standards and hence fulfills the requirements of MDR (EU) 2017/745



A handwritten signature in blue ink, appearing to be 'Ha-deuk Lee', is written over the page.

Ha-deuk Lee / President of Xenosys Co., Ltd.

January 29, 2021