



# EC Declaration of Conformity

Manufacturer:

whose single Authorized Representative:

**Guilin Woodpecker Medical Instrument Co., Ltd.**

**MedNet EC-REP GmbH•Borkstrasse 10  
• 48163 Muenster • Germany**

We, the manufacturer, herewith declare that the products  
**Gutta Percha Obturation Device, GMDN-Code: 44859  
MODEL:Fi-P , Fi-G, Fi-E**

meet the provisions of Directive 93/42/EEC which apply to them.

The medical device has been assigned to class IIa according to Annex IX of the Directive 93/42/EEC. It bears the mark



The product concerned has been manufactured under a quality management system according to Annex II of Directive 93/42/EEC.

Compliance of the designated product with the Directive 93/42/EEC has been assessed and certified by the Notified Body

**TÜV Rheinland LGA Products GmbH  
Tillystraße 2, 90431, Nürnberg, Germany**

Certificate No.: HD 2158053-1  
Issue date: 2021-05-19  
Expiry date: 2024-05-26

following the procedure relating to the EC Declaration of Conformity set out in Annex II of Directive 93/42/EEC.

This Declaration of Conformity covers all medical devices as specified in the product list belonging to this declaration and is only valid in connection with a batch specific Certificate of Compliance for all products concerned bearing the CE mark

The above mentioned declaration of conformity is exclusively under the responsibility of

Company: **Guilin Woodpecker Medical Instrument Co., Ltd.**  
Address: Information Industrial Park, GuiLin National High-Tech Zone, GuiLin, GuangXi,  
541004, P.R.China

陈燕 2022.3.9  
Preparation , Date

王波 2022.3.9  
Review , Date

丁加波 2022.3.9  
Legally binding signature, Location, Function