



Certificate of Analysis

Abbott GmbH
Max-Planck Ring 2
65205 Wiesbaden
Germany

Product Name: Alinity i HAVAb IgG Calibrator
List Number: 08P2601 **SCP/Revision:** 08P26-49-01 / 04E
Lot Number: 39389BE00 **Assay Software Version:** 286_002
Expiration Date: 2023-03-04 **Test Date:** 2022-04-05
Lot Size: 1600

Component Name	List Number	Lot Number	Production Date	Expiration Date
Alinity i HAVAb IgG Calibrator 1	8P26K01	39389BE00K	2022-03-04	2023-03-04

Acceptance Criteria		Specifications	Result	Pass
New Target PC level S/CO	S/CO	1.94 – 2.62	2.45	X

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485:2016 and GMP guidelines as applicable.

The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality Assurance (QA):

J. Kroll

Date:

2022-04-17