



# Certificate of Analysis

Abbott GmbH  
Max-Planck Ring 2  
65205 Wiesbaden  
Germany

**Product Name:** Alinity i HAVAb IgG Controls  
**List Number:** 08P2610 **SCP/Revision:** 08P26-49-10 / 04D  
**Lot Number:** 39160BE00 **Assay Software Version:** 286\_002  
**Expiration Date:** 2023-03-04 **Test Date:** 2022-04-01  
**Lot Size:** 1942

Component Name	List Number	Lot Number	Production Date	Expiration Date
Alinity i HAVAb IgG Positive Control	8P26M01	39160BE00M	2022-03-04	2023-03-04
Alinity i HAVAb IgG Negative Control	8P26L01	39160BE00L	2022-03-04	2023-03-04

Acceptance Criteria		Specifications	Result	Pass
Positive Control	Mean SMP/REF ratio	0.746 – 1.254	0.999	X
Negative Control	Mean S/CO	≤ 0.31	0.11	X

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485:2016 and GMP guidelines as applicable.

The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality Assurance (QA): N. Schubert Date: 04. Apr. 2022