



Certificate of Analysis

Abbott Ireland Diagnostics Division,
Lisnamuck,
Longford,
Co. Longford,
Ireland

Product Name: Alinity i 25-OH Vitamin D Controls

List number: 08P4510
Lot number: 37436UD00
Expiry Date: 2023-04-30

SCP / revision 8P450010/002

Component name	List Number	Lot Number	Production date	Expiry date
Alinity i Vit D 25OH Control L	8P45L	37433UD00	2022-02-01	2023-05-01
Alinity i Vit D 25OH Control M	8P45M	37434UD00	2022-02-01	2023-05-01
Alinity i Vit D 25OH Control H	8P45N	37435UD00	2022-01-31	2023-04-30

Acceptance Table: Mean % Difference Reported Result

Description	Lower Limit	Upper Limit	Value	Result
Low Control	-10	10	-1	Acceptable
Medium Control	-10	10	1	Acceptable
High Control	-10	10	1	Acceptable

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality:

Caimita Fabricia Fawcett
Printed name and signature

Date:

26 Mar 22