



Document Information:

Document Author: Schwaiger, Nicole {DOMS~Penzberg}
Business Area / Unit: Roche Professional Diagnostics
Confidentiality: Confidential
Document Class: Device Master Record
Document Type: Device Master Record
Document Creator: Schwaiger, Nicole {DOMS~Penzberg}
Document Lifecycle Status: Signed
Valid From: 19-Oct-2021 06:52:06 (UTC)
Valid To:
Document Title: C_08_RSP_07212780190_EN
Document Number: 000000000001200000109315
Document Version: 04
Template: No

Global Group: Regulatory Affairs
Global SubGroup: Specification Prod. (Registr.)
Local Group: Reg. Specification
Language: English
Site: RDG Germany
Department: C_DOMS
Document Applies To: C_Standardization/ QC Penzberg
Document Description: Vitamin B12 II CalSet - "Only for use in Regulatory Affairs"

Electronic Signatures:

Signed By: schrankr (Ralf Schrank {DOMSW})
Role: Author
Signature Differentiation: Quality Control
Signed Date: 14-Oct-2021 17:53:54 (UTC)

Signed By: schwaigh (Nicole Schwaiger {DOMSTO})
Role: Reviewer
Signature Differentiation: Quality Control
Signed Date: 19-Oct-2021 05:14:24 (UTC)

Signed By: tassonir (Richard Tassoni {DOMS})
Role: Approver
Signed Date: 19-Oct-2021 06:40:37 (UTC)

SPECIFICATION
for
Vitamin B12 II CalSet

0721 2780 190

Bottle 1 (B₁₂ II Cal1)

Appearance
Vitamin B₁₂

lyophilized substance
approx. 250 pg/mL or 185 pmol/L

Bottle 2 (B₁₂ II Cal2)

Appearance
Vitamin B₁₂

lyophilized substance
approx. 1500 pg/mL or 1107 pmol/L

cobas e 801 analyzer: The exact lot-specific calibrator values are encoded in the electronic barcode and available via the **cobas** link.

All other analyzers: The exact lot-specific calibrator values are encoded in the barcode as well as printed on the enclosed (or electronically available) calibrator barcode sheet.

All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV.

The testing methods use assays that have been approved by the FDA or that are in compliance with the legal rules applicable to placing in vitro diagnostic medical devices for human use on the market in the European Union.