



# Certificate of Analysis

Abbott Ireland Diagnostics Division,  
Lisnamuck,  
Longford,  
Co. Longford,  
Ireland

**Product Name:** Cholesterol2  
**List number:** 04S9220  
**Lot number:** 38245UD00  
**Reagent Lot Number:** 38245UD00  
**Expiry Date:** 2023-07-04

SCP / revision 4S920020/003

Component name	List Number	Lot Number	Production date	Expiry date
Chol2 R1	04S92R	38244UD00	2022-04-04	2023-07-04

## Acceptance Table: Grand Mean Concentration (mg/dL)

Description	Lower Limit	Upper Limit	Value	Result
REF_Control 1	260.94	282.45	274.27	Acceptable
REF_Control 2	99.28	107.29	103.22	Acceptable
REF_GC1_Lv5	581.82	643.06	618.62	Acceptable

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality:

*Kathleen Green*  
Printed name and signature

Date:

16 May 22