

Certificate of Analysis

| | |
|--|--------------------------------------|
| Product Name: TRANSFERRIN CLINICAL CHEMISTRY | Abbott List Number: 01E0442 |
| OEM Part Number: 01E0442 | Abbott Commodity Number: 01E0442 |
| OEM Lot Number: 78395UN21 | Expiration Date: 2022-03-15 |
| Storage: 2 - 8°C | Package Insert No./Rev. G95903 / R03 |
| Qty Manufactured: 179 Kits | |

| Specification | Specification Limit | Result |
|-------------------------|----------------------------------|---------------|
| Barcode Scan | Acceptable verification on c8000 | Complies Pass |
| Bio-Rad L1 | 130 - 159 mg/dL | 144 Pass |
| Bio-Rad L1 | Lot Number | 68941 |
| Bio-Rad L2 | 234 - 286 mg/dL | 264 Pass |
| Bio-Rad L2 | Lot Number | 68942 |
| Bio-Rad L3 | 331 - 405 mg/dL | 371 Pass |
| Bio-Rad L3 | Lot Number | 68943 |
| Control 4: % Difference | -10 - 10 % | 0 Pass |
| Micro: Beg Sample (R1) | <30 CFU/mL | 1 Pass |
| Micro: End Sample (R1) | <30 CFU/mL | 0 Pass |
| Micro: Beg Sample (R2) | <30 CFU/mL | 0 Pass |
| Micro: End Sample (R2) | <30 CFU/mL | 0 Pass |

For in-vitro diagnostic use.

This lot meets all current release specifications per quality document MS.MASTER342.

The above product has been manufactured in compliance with Quality Systems Regulations, passed all Thermo Fisher final release specifications, and has met all Abbott acceptance requirements.

This product contains animal source material. The caprine material used in the manufacture of this product is derived from goat antiserum collected from U.S. sourced goats.

This product contains R1 and R2 components. The component name, lot number, and expiration date matches the kit name, lot number, and expiration date listed on this COA.



Quality Representative

Date: 2021-03-29