Certificate of Analysis

Abbott Ireland Diagnostics Division,
Lisnamuck,
Longford,
Co. Longford,
Ireland

Product Name: Albumin BCP2 SCP / revision 4U450030/005
List number: 04U4530
Lot number: 40184UD00
Reagent Lot Number: 40184UD00
Expiry Date: 2023-07-13

Component name	List Number	Lot Number	Production date	Expiry date
AlbBCP2 R1	04U45R	40182UD00	2022-04-13	2023-07-13

Acceptance Table: Grand Mean Concentration (g/dL)

Description	Lower Limit	Upper Limit	Value	Result
REF_Control_1	3.70	4.19	3.98	Acceptable
REF_Control_2	2.37	2.81	2.64	Acceptable
REF_GC1_Lv5	6.42	7.52	6.82	Acceptable

Cruz Bella Rose Navarro

Digitally signed by Cruz Bella Rose Navarro
DN: c=VN, st=THÀNH PHỐ HÀ NỘI, o=Văn phòng đại diện ABBOTT LABORATORIES GMBH tại thành phố Hà Nội, title=Trưởng Đại Diện, cn=Cruz Bella Rose Navarro, 0.9.2342.19200300.100.1.1=MST:P5954425A
Date: 2022.06.10 10:10:09 +07'00'

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality: Date: 05 May 22
Printed name and signature

Abbott
A Promise for Life