



# Certificate of Analysis

Abbott Ireland Diagnostics Division,  
Lisnamuck,  
Longford,  
Co. Longford,  
Ireland

**Product Name:** Albumin BCG2      SCP / revision      4U300020/004  
**List number:** 04U3020  
**Lot number:** 38492UD00  
**Reagent Lot Number:** 38492UD00  
**Expiry Date:** 2023-05-24

Component name	List Number	Lot Number	Production date	Expiry date
AlbBCG2 R1	04U30R	38491UD00	2022-02-24	2023-05-24

## Acceptance Table: Grand Mean Concentration (g/dL)

Description	Lower Limit	Upper Limit	Value	Result
REF Control 1	4.07	4.63	4.27	Acceptable
REF Control 2	2.58	3.04	2.76	Acceptable
REF GC1 Lv5	6.65	7.79	7.21	Acceptable

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality:

*Kathleen Anne Victoria Grech*  
Printed name and signature

Date:

*08 Mar 22*