



Certificate of Analysis

Abbott Ireland Diagnostics Division,
Lisnamuck,
Longford,
Co. Longford,
Ireland

Product Name: Architect Free T4 Calibrators

List number: 7K65-02

Lot number: 39038UD00

Expiry Date: 2023-09-03

SCP / revision

7K650002/011

Component name	List Number	Lot Number	Production date	Expiry date
Architect Free T4 Calibrator A	7K65A	39032UD00	2022-03-03	2023-09-03
Architect Free T4 Calibrator B	7K65B	39033UD00	2022-03-03	2023-09-03
Architect Free T4 Calibrator C	7K65C	39034UD00	2022-03-04	2023-09-04
Architect Free T4 Calibrator D	7K65D	39035UD00	2022-03-04	2023-09-04
Architect Free T4 Calibrator E	7K65E	39036UD00	2022-03-04	2023-09-04
Architect Free T4 Calibrator F	7K65F	39037UD00	2022-03-04	2023-09-04

Acceptance Table: Grand Mean (n=4) Concentration (ng/dL) over 2 instruments

Description	Lower Limit	Upper Limit	Value	Result
REF Control Low (ng/dL)	0.57	0.72	0.66	Acceptable
REF Control Medium (ng/dL)	1.06	1.42	1.21	Acceptable
REF Control High (ng/dL)	2.19	2.96	2.76	Acceptable
REF Panel N (ng/dL)	1.04	1.39	1.12	Acceptable

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality:

CARMA FERRI (C Ferreri)
Printed name and signature

Date:

04 April 22