



# Certificate of Analysis

Abbott Ireland Diagnostics Division,  
Lisnamuck,  
Longford,  
Co. Longford,  
Ireland

**Product Name:** Architect Free T4 Controls

**List number:** 7K65-10

**Lot number:** 37398UD00

**Expiry Date:** 2023-05-02

SCP / revision

7K650010/013

Component name	List Number	Lot Number	Production date	Expiry date
Architect Free T4 Control Low	7K65L	37395UD00	2022-03-02	2023-05-02
Architect Free T4 Control Medium	7K65M	37396UD00	2022-03-02	2023-05-02
Architect Free T4 Control High	7K65N	37397UD00	2022-03-02	2023-05-02

## Acceptance Table: Grand Mean (n=4) Conc (ng/dL) over 2 Instruments

Description	Lower Limit	Upper Limit	Value	Result
SMP Control Low (n=4) (ng/dL)	0.56	0.72	0.63	Acceptable
SMP Control Medium (n=4) (ng/dL)	1.01	1.47	1.21	Acceptable
SMP Control High (n=4) (ng/dL)	2.18	3.29	2.90	Acceptable

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality:

Date:

Caroline Hayes C-HAYES  
Printed name and signature

Date:

04 April 2022