



# Certificate of Analysis

Abbott Ireland Diagnostics Division,  
Lisnamuck,  
Longford,  
Co. Longford,  
Ireland

**Product Name:** ARCHITECT Free T4 Reagent Kit  
**List number:** 7K65-29  
**Lot number:** 40003UD01  
**Reagent Lot number:** 40003UD00  
**Expiry Date:** 2023-04-20

SCP / revision 7K650029/003

Component name	List Number	Lot Number	Production date	Expiry date
Architect Free T4 Microparticles	7K65G	40001UD00	2022-04-20	2023-04-20
Architect Free T4 Conjugate	7K65H	40002UD00	2022-05-17	2023-10-17

## Acceptance Table: Grand Mean (n=4) Concentration (ng/dL) over 2 Instruments

Description	Lower Limit	Upper Limit	Value	Result
REF Control Low	0.57	0.72	0.67	Acceptable
REF Control Medium	1.05	1.47	1.22	Acceptable
REF Control High	2.18	3.31	2.80	Acceptable
REF Panel N	0.98	1.39	1.14	Acceptable

This product has been manufactured, including packaging and quality control in full compliance with ISO 13485, IVD Directive 98/79/EC and GMP guidelines as applicable.

Note: The specifications in this certificate reflect the internal Abbott release specifications. When performing the assay in a laboratory the specifications listed in the package insert apply.

Quality:

  
Printed name and signature

Date:

28 May 22