

# BioSystems

## Quality Standard Procedures

### **ADA Standard**

All test procedures are carried out in BioSystems Clinical Chemistry Analyzer as a Quality Control Department routine.

#### *Control of the assigned value*

Measurement of the concentration of an analyte against a standard/calibrator of a previous batch.

Express the result as % difference from the assigned value.

Tolerance: + / - 10%

Expanded Uncertainty < 5%



BioSystems S.A.

June 2021