

BioSystems

Quality Standard Procedures

FERRITIN Standard

All test procedures are carried out in BioSystems Clinical Chemistry Analyzer as a Quality Control Department routine.

Control of the assigned value

Measurement of the concentration of an analyte against a standard/calibrator of a previous batch.

Express the result as % difference from the assigned value.

Tolerance: + / - 10%

Expanded Uncertainty < 5%



BioSystems S.A.

June 2021