HEMATOCRIT CONTROL KIT

SCLBL1366

For use with

Blood Gas Analyzers Only.

Intended Use

Hematocrit Controls (Level-1 & Level-2) used for monitoring of Hematocrit Level on Blood Gas Analyzer. Sensa Cores Hematocrit Controls (Level-1 & Level-2) is not made for examine specimens from the human body for diagnostic purposes and made for clinical laboratory equipment use only.

Pack Size:

Hct Level - 1 (**14 - 34%**) : 6 Bottles (6 x 4 mL)

Hct Level - 2 (**35 - 55%**) : 6 Bottles (6 x 4 mL)

Storage and Stability

Store at 18-25 °C until expiration date mentioned on labels

CE	CE Marking	IVD	In vitro diagnostic medical device
	Manufacturer		Caution
	Date of Manufacture		Do Not Use if Package is Damaged
	Use by Date	ISO	ISO Certified
LOT	Batch Code		Keep Dry
	Keep Away from Sunlight		Do not discard along with household waste
	Storage & Operating Temperature		Non Flammable
	Consult instructions for use		Keep away from Flame

Irritant! Irritating to eyes and skin. Avoid contact with skin and eyes. In case of contact with eyes, rinse immediately with plenty of water and seek medical advice.

Sensa Core Medical Instrumentation Pvt. Ltd.
Plot No: 3, Export Promotion Industrial Park, IDA, Pashamylaram,
Sangareddy (District) - 502307, Hyderabad, Telangana, INDIA.
Tel: +91 8455 - 223400 www.sensacore.com

EC REP

ISO 9001 ISO 13485