

# SENSA STAT CONTROL KIT

## For use with

SCLBL1672

**Blood Gas Analyzers Only.**

## **Intended Use**

Sensa Stat Controls (Level-1, Level-2 & Level-3) used for monitoring of Sodium, Potassium, Ionized Calcium, pH, Chloride, Glucose, Lactate, pCO<sub>2</sub>, pO<sub>2</sub> on the Blood Gas Analyzer. Sensa Cores Sensa Stat Controls (Level-1, Level-2 & Level-3) is not made for examine specimens from the human body for diagnostic purposes and made for clinical laboratory equipment use only. Sensa Cores Sensa Stat Controls (Level-1, Level-2 & Level-3) is Single use only.

## Pack Size:




















**Sensa Stat Level - 1** : 10 Ampules (10 x 1.8 mL)

**Sensa Stat Level - 2** : 10 Ampules (10 x 1.8 mL)

**Sensa Stat Level - 3** : 10 Ampules (10 x 1.8 mL)

## **Storage and Stability**

Store at 18-25°C until expiration date mentioned on labels

 CE	CE Marking	 IVD	In vitro diagnostic medical device
 	Manufacturer		Caution
 	Date of Manufacture		Do Not Use if Package is Damaged
 	Use by Date		ISO Certified
 LOT	Batch Code		Keep Dry
	Keep Away from Sunlight		Do not discard along with household waste
	Storage & Operating Temperature		Non Flammable
	Consult instructions for use		Keep away from Flame

***Irritant!*** Irritating to eyes and skin. Avoid contact with skin and eyes. In case of contact with eyes, rinse immediately with plenty of water and seek medical advice.

 **Sensa Core Medical Instrumentation Pvt. Ltd.**  
Plot No: 3, Export Promotion Industrial Park, IDA, Pashamylaram,  
Sangareddy (District) - 502307, Hyderabad, Telangana, INDIA.  
Tel: +91 8455 - 223400 www.sensacore.com

EC REP

ISO 9001  
ISO 13485

ISO 9001

ISO 13485