

**BACT/ALERT® BPA**

IVD

**Intended Use**

**BACT/ALERT® BPA** culture bottles are used with BACT/ALERT® Microbial Detection Systems (BACT/ALERT® 3D and BACT/ALERT® VIRTUO®) for quality control testing of leukocyte-reduced apheresis platelet (LRAP) units, both single and pools of up to six (6) units of leukocyte-reduced whole blood platelet concentrates (LRWBPC), and pools of up to four (4) units of leukocyte-reduced whole blood derived buffy coat platelets (LRWBDBCP).<sup>1,2</sup> BACT/ALERT® BPA culture bottles support the growth of aerobic microorganisms (bacteria and fungi).

**Summary and Explanation**

BACT/ALERT® Microbial Detection Systems provide both a microbial detection system and a culture media with suitable nutritional and environmental conditions for organisms which might be present in the test sample. Inoculated bottles are placed into the instrument where they are incubated and continuously monitored for the presence of microorganisms that will grow in the BACT/ALERT® BPA bottles.

BACT/ALERT® Microbial Detection Systems may be used for quality control testing of platelets. The laboratory should follow its own quality control procedures for this use. BACT/ALERT® Microbial Detection Systems, including the culture bottles, were not cleared for use in determining suitability for release of platelets for transfusion. Users considering such release testing should first consult the national regulatory agency for requirements and studies necessary to support that release testing.

The performance of BACT/ALERT® Microbial Detection Systems for the detection of bacteria in non-leukocyte-reduced platelet products is not known since studies were conducted utilizing LRAP, and leukocyte-reduced WBPC and WBDBCP products. The testing of platelets and non-leukocyte-reduced platelet products alone should not be used to extend the shelf life of platelets without consulting your national regulatory agency.

**Note:** The information provided applies to all configurations of BACT/ALERT® Microbial Detection Systems, unless otherwise noted.

**Principle of the Test**

BACT/ALERT® Microbial Detection Systems utilize a colorimetric sensor and reflected light to monitor the presence and production of carbon dioxide (CO<sub>2</sub>) that is dissolved in the culture medium. If microorganisms are present in the test sample, carbon dioxide is produced as the organisms metabolize the substrates in the culture medium. When growth of the microorganisms produces CO<sub>2</sub>, the color of the gas-permeable sensor installed in the bottom of each culture bottle changes to yellow.<sup>3</sup>

<sup>1</sup> Brecher ME, Hay SN, Rose AD, Rothenberg SJ. Evaluation of BacT/ALERT plastic culture bottles for use in testing of pooled whole blood derived leukocyte-reduced PRP platelets with a single contaminated unit. *Transfusion*, 2005; 45: 1512-1517.

<sup>2</sup> Brecher ME, Hay SN, Rothenberg SJ. Evaluation of a new generation of plastic culture bottle with an automated microbial detection system for nine common contaminating organisms found in platelet components. *Transfusion* 2004; 44: 359-363.

<sup>3</sup> Thorpe TC, Wilson ML, Turner JE, et al. BacT/ALERT: an Automated Colorimetric Microbial Detection System. *J Clin Micro* 1990; 28 (7), 1608-1612.

---

## Reagents

---

For *in vitro* diagnostic use only.

**Caution:** Handle specimens and inoculated culture bottles as though capable of transmitting infectious agents. All inoculated culture bottles and specimen collection needles should be decontaminated according to your institution's procedures.<sup>4</sup>

**BACT/ALERT® BPA** (color-coded blue) – BACT/ALERT® BPA disposable culture bottles contain 40 mL of media and an internal sensor that detects carbon dioxide as an indicator of microbial growth. The BACT/ALERT® BPA culture bottle does not require venting. The media formulation consists of pancreatic digest of casein (1.7% w/v), papaic digest of soybean meal (0.3% w/v), sodium polyanetholesulfonate (SPS) (0.035% w/v), pyridoxine HCl (0.001% w/v), and other complex amino acid and carbohydrate substrates in purified water. Bottles are prepared with an atmosphere of CO<sub>2</sub> in oxygen under vacuum. The composition of the media may be adjusted to meet specific performance requirements.

**Caution:** The BACT/ALERT® plastic bottles contain polycarbonate. Since not all disinfectants are intended for use with polycarbonate surfaces, please refer to the product labeling of the disinfectant to verify compatibility.

**Caution:** Platelet specimens determined positive by BACT/ALERT® may contain organisms that are positive by smear that will not grow on routine subculturing media. These specimens should be subcultured on special media when such organisms are suspected. Also, BACT/ALERT® positive specimens may contain organisms that are not seen with routine smear methods and may require both specialized smears and subculturing media for detection and recovery.

**Caution:** On rare occasions, organisms may be encountered that grow in the BACT/ALERT® BPA culture bottle growth medium but do not produce sufficient carbon dioxide to be determined positive. At maximum test time the bottle should be visually inspected for evidence of growth.

**Caution:** Prompt removal of positives as they are signaled by BACT/ALERT® is strongly recommended to avoid possible non-viable cultures due to autolysis or other reasons. Certain strains of *Streptococcus pneumoniae* may be particularly prone to autolysis if they are not removed promptly after being signaled positive.

### Additional Materials Required

- BACT/ALERT® Microbial Detection Systems
- Sterile Airway Needle/Subculture Units
- Disposable gloves
- Appropriate biohazard waste containers for materials potentially contaminated with infectious agents
- Appropriate platelet coupling and sampling apparatus or platelet bag with integrally connected sample bag
- Alcohol pads or equivalent

### Materials Available from bioMérieux

- BACT/ALERT® Microbial Detection Systems
- Sterile Airway Needle/Subculture Units

### Storage Instructions

BACT/ALERT® BPA culture bottles are ready for use. Store in an upright position protected from direct light at room temperature (15-30°C). An expiration date is printed on each bottle label. Do not inoculate the culture bottles beyond the expiration date indicated. If the bottles are exposed to temperatures less than 15°C, precipitates may form that will disappear when the bottles are warmed to room temperature. Bottles must be at room temperature before use.

### Chemical or Physical Indications of Instability

Prior to use, the BACT/ALERT® BPA culture bottles should be examined for evidence of damage or deterioration (discoloration). Bottles exhibiting evidence of damage, leakage, or deterioration should be discarded. The medium in undisturbed bottles should be clear, but there may be a slight opalescence or a trace of precipitate due to the anticoagulant SPS; do not confuse this with turbidity. Do not use a bottle which contains medium exhibiting turbidity, a yellow sensor, or excess gas pressure; these are signs of possible contamination.

---

<sup>4</sup> Widmer AF, Frei R. Decontamination, Disinfection, and Sterilization, in Murray PR (ed.). *Manual of Clinical Microbiology*, ed. 7. Washington, D.C., American Society for Microbiology, 1999, pp 138-164.

---

## Instruments

---

Review the appropriate BACT/ALERT® Microbial Detection System User Manual before use.

## Specimen Collection and Preparation

---

The leukocyte-reduced platelet specimen must be collected using sterile procedures such that the collection set remains a closed system (e.g., use of an integrally connected sample bag or a sample bag connected with a sterile connection device, such as a tubing welder, per the device manufacturer's instructions). It is recommended to use disposable gloves when handling the sampling site and sampling bag to reduce the risk of contaminating the sampling site and sampling site coupler. Refer to Cumitech 1C<sup>5</sup> for the proper contamination avoidance procedure. The platelet specimen should be taken at least 24 hours after collection to allow for natural proliferation in the platelet product.<sup>6</sup>

**Note:** Some literature suggests that the sampling of platelet product for bacterial contamination has shown to be most sensitive when taken no sooner than 36 hours or 48 hours following collection/pooling, depending on the expected shelf-life of the platelets. Consult local regulatory guidance for further requirements. Minimum sampling delay times should be validated as part of the platelet testing program.<sup>7,8</sup>

General suggested guidelines for preparing and collecting the platelet specimen for testing are provided below.

1. Label the sample bag with the platelet product information.
2. The platelet specimen to be tested should be taken from the platelet bag(s) using an integrated sampling bag or sterile sampling device. If the platelet bag does not have an integrated sampling bag, a sterile connection device, such as a tubing welder, should be used to connect a sterile sampling bag or device in order to preserve the integrity of the platelet product, so that a closed system is maintained.
3. Strip the attached tubing between the platelet bag (LRAPs, single LRWBPC or a pool of up to 6 units of LRWBPC or a pool of up to 4 units of LRWBDBCP) and the sample bag toward the platelet bag, rotate contents of platelet bag to allow thorough mixing, and allow the tubing to refill from the platelet bag. Repeat an additional two times. Fill the sample bag with volume desired. Heat seal the tubing between the platelet bag and the sample bag. Aseptically remove the sample bag by cutting the tubing between two of the heat seal welds.

**Note:** A sample volume of 4-10 mL is required for each culture bottle to be inoculated.

4. For single bag sampling of a unit of whole blood platelet concentrate or sampling of a pool of up to 6 units of whole blood platelet concentrates or sampling of a pool of up to 4 units of whole blood derived buffy coat platelets, use an integrally connected sterile sample bag or a sample bag that has been attached using a sterile connection device, such as a tubing welder. Remove the desired test volume to the sample bag. Seal the sample bag off from the platelet bag, separate, and inoculate culture bottles from the sample bags.

---

## BACT/ALERT® BPA Culture Bottle Test Procedure

---

### Preliminary Comments and Precautions

1. For best overall recovery when culturing platelet specimens, it is strongly recommended that more than one type of culture bottle be utilized (e.g., one aerobic and one anaerobic).
2. DO NOT VENT BACT/ALERT® BPA BOTTLES. Positive culture bottles should be transiently vented before subculturing, staining, or disposal to release any gas produced during microbial metabolism.
3. Use disposable gloves and handle inoculated bottles cautiously as though capable of transmitting infectious agents. Consult a physician immediately if contaminated materials are ingested or come in contact with open lacerations, lesions, or other breaks in skin.

---

<sup>5</sup> Baron EJ, Weinstein MP, Dunne WM Jr, Yagupsky P, Welch DF, Wilson DM. 2005. Cumitech 1C, *Blood Cultures IV*. Coordinating ed., Baron EJ. ASM Press, Washington, DC.

<sup>6</sup> Brecher ME, Holland PV, Pineda AA, Tegtmeier GE, Yomtovian R. Growth of bacteria in inoculated platelets: implications for bacteria detection and the extension of platelet storage. *Transfusion* 2000; 40:1308-1312.

<sup>7</sup> Brecher ME, Holland PV, Pineda AA, Tegtmeier GE, Yomtovian R. Growth of bacteria in inoculated platelets: implications for bacteria detection and the extension of platelet storage. *Transfusion* 2000; 40:1308-1312.

<sup>8</sup> McDonald C, Allen J, Brailsford S, et al. Bacterial screening of platelet components by National Health Service Blood and Transplant, an effective risk reduction measure. *Transfusion* 2017; 57 (5), 1122–1131.

4. When handling positive bottles that are bulging or leaking, wear appropriate personal protective equipment (PPE) to avoid coming in contact with microorganisms.
5. Immediately clean up any spillage of contaminated material using a 1:10 dilution of 5% sodium hypochlorite. Dispose of the cleaning material by an acceptable method.
6. All inoculated culture bottles and specimen collection needles should be decontaminated according to your institution's procedures.<sup>9</sup>
7. Culture bottles should be utilized by trained laboratory personnel.

#### Procedural Notes and Precautions

1. Great care must be taken to prevent contamination of the platelet sample during inoculation into the culture bottles. Contamination could lead to a specimen being determined positive when a clinically relevant isolate is not actually present in the donated platelet unit.

**Note:** When sampling platelets, it should not be assumed that a sampling error leads to a positive culture of common skin contaminants (e.g., *Staphylococcus aureus*, *Staphylococcus epidermidis*).<sup>10</sup>

2. If inoculated culture bottles have been delayed in their receipt into the lab or have been incubated prior to entry into the BACT/ALERT® instrument, they should be visually inspected for indications of microbial growth. If microbial growth is evident, treat the bottles as positive and do not place in the BACT/ALERT® Microbial Detection System for monitoring.
3. Likely causes of contamination can occur from inadequate aseptic/sterile technique or operator error (e.g., operator lab coat, aerosol), sampling or inoculation in an inadequate environment, or a spore present on top of the BACT/ALERT® bottle septum when introducing the specimen which was not removed with the 70% alcohol wipe.

## Specimen Test/Inoculation Procedure

### Platelet Test Procedure

1. Label the culture bottles with the platelet product information. The bottle must be at room temperature.
2. Remove the plastic flip-top from each culture bottle and disinfect the septum with an alcohol pad or equivalent. Allow to air dry.
3. Disinfect the rubber septum on the surface of the platelet bag sampling site with an alcohol pad or equivalent, allow to air dry, and use a syringe and needle (using a needle gauge sufficiently large enough to allow easy drawback of platelet product into the syringe) to remove a sample from the sample bag. Alternatively, a sterile, integrally connected sampling device may be used to obtain a sample from the platelet bag.

**Note:** A sample volume of 4-10 mL is required for each culture bottle to be inoculated.

4. Insert the needle through the septum of the culture bottle and inject 4-10 mL of the platelet specimen into each bottle being inoculated. If using both an anaerobic and aerobic culture bottle, transfer to the anaerobic bottle first, so that any oxygen trapped in the syringe will not be transferred to this bottle. If a sterile, integrally connected sampling device is used, then the aerobic bottle must be inoculated first, followed by the anaerobic bottle, in order to minimize transfer of oxygen to the anaerobic bottle.

**Caution:** Never force the syringe plunger down during inoculation, as splashing of sample may occur. Remove the syringe when the fill amount is reached, as the vacuum will automatically draw more than the recommended maximum. Puncture the bottle stopper vertically to avoid releasing the vacuum; a bottle without a vacuum should not be inoculated.

5. Ensure that the specimen is properly mixed with the reagents in the BACT/ALERT® BPA bottle.

### Laboratory Procedure

1. DO NOT VENT BACT/ALERT® BPA BOTTLES.
2. Visually inspect bottles before testing. Bottles with a yellow sensor, turbidity, excess gas pressure, and/or evidence of growth should be treated as positive. Smear and subculture. Do not incubate unless smear is negative.
3. After collection, promptly transport the inoculated bottle to testing laboratory and test immediately.
4. After culture bottles have been loaded into the instrument, they should remain there for five to seven days or until designated positive, or until the platelet unit reaches its expiration date.

<sup>9</sup> Widmer AF, Frei R. Decontamination, Disinfection, and Sterilization, in Murray PR (ed.). *Manual of Clinical Microbiology*, ed. 7. Washington, D.C., American Society for Microbiology, 1999, pp 138-164.

<sup>10</sup> Sazama K. Bacteria in blood for transfusion. *Arch Pathol Lab Med*; April 1994; 118:359.

5. All bottles designated positive should be smeared and subcultured. If the smear is negative, indicating a possible false positive, the bottle should be reloaded into the instrument until growth of subculture or redesignation as positive. Cultures which were initially determined false positive and were redesignated positive should be smeared and subcultured.
6. Negative cultures may be checked by smear and/or subculture at some point prior to discarding as negative.
7. Procedures for loading and unloading culture bottles into the appropriate BACT/ALERT® instrument are given in the User Manual.
8. **Do not reuse BACT/ALERT® culture bottles.** Dispose of inoculated BACT/ALERT® culture bottles according to your laboratory protocol. Autoclaving and/or incinerating inoculated BACT/ALERT® bottles is appropriate.<sup>11</sup>

**Note:** A report of “negative” should not be interpreted as meaning that the original product is sterile. The negative status could be due to under-inoculation of the bottle, no organisms present in the inoculum, the number of organisms were too small for detection, or a culture bottle/medium that does not support the growth of the organism. For best overall recovery when culturing platelet specimens it is strongly recommended that more than one type of culture bottle be utilized (e.g., one aerobic and one anaerobic).

**Note:** For quality control testing of platelets, the culture bottle should be held through the expiration date of the product (The platelet unit with the shortest expiration date in the 6 unit pool will determine the final expiration of the pool.) or until designated positive. The platelet specimen should be taken at least 24 hours after collection to allow for natural proliferation in the platelet product.<sup>12</sup>

---

## Quality Control

A Certificate of Conformance is available for each lot of culture bottles. If desired, individual laboratories can perform quality control testing of BACT/ALERT® BPA culture bottles. Refer to the appropriate BACT/ALERT® User Manual and to CLSI® document M22-A3.<sup>13</sup>

### Instrument

A BACT/ALERT® Reflectance Standards kit is provided with each instrument for the QC and Calibration procedures. All quality control should be part of normal system maintenance. Refer to the appropriate BACT/ALERT® User Manual for more information.

**Caution:** If your facility's LIS vendor sends bottle IDs and bottle type abbreviations to the BACT/ALERT® instrument, use the correct bottle type abbreviation to avoid possible false positive or false negative results. For more information, contact your local bioMérieux representative.

---

## Results

Positive or negative culture bottles are determined by decision-making software contained in the BACT/ALERT® Microbial Detection Systems. No action is required until the BACT/ALERT® instrument signals culture bottles positive or negative.

---

## Limitations of the Test

Many variables involved in platelet culture testing cannot be practically controlled to provide total confidence that results obtained are due solely to proper or improper performance of any culture medium or detection system.

1. A Gram-stained smear from a negative bottle may sometimes contain a small number of non-viable organisms that were derived from culture medium components, staining reagents, immersion oil, or glass slides, therefore, false-positive results are indicated.
2. False positive readings can occur due to noise on the powerline, placing the instrument in direct sunlight, or with dramatic temperature fluctuations.

---

<sup>11</sup> Brecher ME, Hay SN, Rothenberg SJ. Evaluation of a new generation of plastic culture bottle with an automated microbial detection system for nine common contaminating organisms found in platelet components. *Transfusion* 2004; 44: 359-363.

<sup>12</sup> Brecher ME, Holland PV, Pineda AA, Tegtmeier GE, Yomtovian R. Growth of bacteria in inoculated platelets: implications for bacteria detection and the extension of platelet storage. *Transfusion* 2000; 40:1308-1312.

<sup>13</sup> CLSI®/NCCLS. *Quality Control for Commercially Prepared Microbiological Culture Media*; Approved Standard—Third Edition. CLSI®/NCCLS document M22-A3. Wayne, PA: NCCLS; 2004.

3. Failure to achieve adequate leukocyte reduction may result in false positive readings.

## Performance Characteristics

### BACT/ALERT® 3D Microbial Detection Systems

#### Detection of Organisms in Leukocyte-Reduced Apheresis Platelets

A study to determine the ability of the culture bottles to detect the presence of microorganisms in leukocyte-reduced apheresis platelets was performed at one clinical site. Bags were seeded at Day 2 with nine individual microorganisms to include:

- *Escherichia coli* ATCC® 25922™
- *Enterobacter cloacae* clinical isolate
- *Klebsiella oxytoca* clinical isolate
- *Cutibacterium acnes* clinical isolate
- *Serratia marcescens* ATCC® 43862™
- *Staphylococcus aureus* ATCC® 27217™
- *Staphylococcus epidermidis* ATCC® 49134™
- *Streptococcus viridans* group clinical isolate

Three replicates of each bottle type were inoculated (4 mL) with each organism at each inoculum level. Seventy-two bottles at one site were inoculated with 4 mL of platelets (no seeded organisms) to serve as negative controls at the lower sample volume range, i.e., 4 mL, and 408 bottles at two sites were inoculated with 10 mL of platelets (no seeded organisms) to serve as negative controls of the higher sample volume range, i.e., 10 mL. There were no false positives from the negative controls inoculated at 4 mL and one false positive from the negative controls inoculated at 10 mL (1/408 or 0.25%). The initial concentration of organisms seeded varied for each organism and ranged from 1 CFU/mL to 300 CFU/mL. See the following table for results.

**Table 1: Recovery of Organisms in Leukocyte-Reduced Apheresis Platelets**

Microorganism	ATCC® Number	Actual Initial Inoculum (CFU/mL)*	Average Time to Detection (hours)			
			Plastic BACT/ALERT® BPA†	Glass BACT/ALERT® SA†	Plastic BACT/ALERT® BPN†	Glass BACT/ALERT® SN†
<i>Bacillus cereus</i>	11778™	5	8.7 (8.7-8.8)	8.9 (8.8-9.0)	9.7 (9.5-9.8)	10.4 (9.7-10.7)
		2	9.3 (9.1-9.5)	9.7 (9.3-10.1)	10.9 (10.7-11.1)	10.8 (10.2-11.1)
	25922™	215	9.9 (9.7-10.1)	10.8 (10.6-11.1)	9.3 (9.1-9.4)	10.2 (10.1-10.3)
		5	11.1 (10.9-11.2)	12.0 (11.9-12.1)	10.3 (10.2-10.4)	11.0 (10.9-11.1)
<i>Enterobacter cloacae</i>	Clinical Isolate	300	9.9 (9.8-10.0)	10.7 (10.7-10.8)	9.6 (9.5-9.7)	10.5 (10.4-10.5)
		21	11.0 (10.9-11.2)	11.8 (11.7-11.8)	10.3 (10.2-10.4)	11.5 (11.4-11.5)
	Clinical Isolate	32	10.1 (9.9-10.3)	10.8 (10.8-10.8)	10.4 (10.3-10.5)	11.2 (11.1-11.3)
		5	10.9 (10.7-11.0)	11.5 (11.5-11.5)	11.1 (11.0-11.2)	12.1 (12.0-12.2)

Microorganism	ATCC® Number	Actual Initial Inoculum (CFU/mL)*	Average Time to Detection (hours)			
			Plastic BACT/ALERT® BPA†	Glass BACT/ALERT® SA†	Plastic BACT/ALERT® BPN†	Glass BACT/ALERT® SN†
<i>Cutibacterium acnes</i>	Clinical Isolate	130	Negative‡	Negative‡	64.0 (62.4-64.8)	80.8 (76.8-86.4)
		16	Negative‡	Negative‡	72.8 (72.0-74.4)	90.4 (88.8-93.6)
<i>Serratia marcescens</i>	43862™	50	11.4 (11.1-11.6)	11.9 (11.8-12.1)	11.8 (11.8-11.9)	12.8 (12.8-12.8)
		5	12.6 (12.4-12.8)	13.0 (12.8-13.3)	12.8 (12.8-12.8)	13.8 (13.8-13.9)
<i>Staphylococcus aureus</i>	27217™	140	10.4 (10.3-10.5)	11.0 (10.8-11.3)	11.5 (11.2-11.8)	12.6 (12.3-12.8)
		4	11.4 (11.1-11.7)	12.0 (12.0-12.1)	12.9 (12.7-13.0)	13.5 (13.0-14.4)
<i>Staphylococcus epidermidis</i>	49134™	40	15.6 (15.5-15.7)	16.9 (16.8-17.0)	19.6 (19.5-19.7)	36.9 (19.3-47.8)
		1	17.3 (17.0-17.5)	18.9 (18.8-19.2)	21.6 (21.3-21.8)	40.1 (23.7-50.4)
<i>Streptococcus viridans</i>	Clinical Isolate	110	15.3 (15.1-15.4)	18.9 (18.6-19.1)	14.7 (14.0-15.2)	18.2 (17.4-19.1)
		3	17.9 (17.2-18.4)	22.4 (21.7-23.2)	18.0 (17.6-18.4)	21.4 (20.7-21.7)
Mean (n = 24 for each aerobic bottle type and n = 27 for each anaerobic bottle type at each inoculum level)		114	11.4§ (8.7-15.7)	12.5§ (8.8-19.1)	17.8 (9.1-64.8)	22.6 (9.7-86.4)
		7	12.7§ (9.1-18.4)	13.9§ (9.3-23.2)	20.1 (10.2-74.4)	25.0 (10.2-93.6)

\* Prior to inoculation, two samples from each platelet bag were inoculated (4 mL per bottle) into plastic BACT/ALERT® BPA and glass BACT/ALERT® SA and plastic BACT/ALERT® BPN and glass BACT/ALERT® SN culture bottles (a total of 72 samples) to verify sterility of the apheresis bags (negative controls). These negative controls were negative, i.e., sterile. In addition, 204 bottles each of plastic BACT/ALERT® BPA and BACT/ALERT® BPN (408 bottles total) were inoculated with 10 mL of non-seeded, sterile, leukocyte-reduced apheresis platelets to serve as additional negative controls. There was one false positive result (0.25%).

† Three replicates of each bottle type were inoculated with each organism at each inoculum level. The average value is listed, with the range of values obtained listed in parenthesis below the average.

‡ High oxygen levels in this bottle prevent the growth of this organism. Recovery occurred in the BACT/ALERT® BPN and SN bottles. **Both BACT/ALERT® BPA and BACT/ALERT® BPN culture bottles should be inoculated from the platelet specimen for optimal recovery of contaminating microorganisms.**

§ Does not include data for *Cutibacterium acnes*.

#### Detection of Organisms in Leukocyte-Reduced Single Units of Whole Blood Platelet Concentrates

A study to determine the ability of the culture bottles to detect the presence of microorganisms in leukocyte-reduced single units of whole blood platelet concentrates was performed. Platelet bags were seeded at Day 2 with nine individual microorganisms to include:

- *Bacillus cereus* ATCC® 11778™
- *Escherichia coli* ATCC® 25922™

- *Enterobacter cloacae* clinical isolate
- *Klebsiella pneumoniae* clinical isolate
- *Cutibacterium acnes* clinical isolate
- *Serratia marcescens* ATCC® 43862™
- *Staphylococcus aureus* ATCC® 27217™
- *Staphylococcus epidermidis* ATCC® 49134™
- *Streptococcus viridans* group clinical isolate

Five replicates of each bottle type were inoculated (4 mL) with each organism at each inoculum level. The initial concentration of organisms seeded varied for each organism and ranged from <2 CFU/mL to 265 CFU/mL. An additional 180 bottles served as negative controls (WBPC with no seeded organisms added). No false positives or contaminated negative controls were detected. See the following table for results.

**Table 2: Recovery of Organisms in Leukocyte-Reduced Single Units of Whole Blood Platelet Concentrates**

Microorganism	ATCC® Number	Actual Initial Inoculum (CFU/mL)*	Number of Positive Cultures		
			BACT/ALERT® BPA†	BACT/ALERT® BPN†	Solid Media
<i>Bacillus cereus</i>	11778™	85	5	5	5
		5	5	5	5
<i>Escherichia coli</i>	25922™	110	5	5	5
		6	5	5	5
<i>Enterobacter cloacae</i>	Clinical Isolate	265	5	5	5
		17	5	5	5
<i>Klebsiella pneumoniae</i>	Clinical Isolate	20	5	5	5
		2	5	5	4
<i>Cutibacterium acnes</i>	Clinical Isolate	28	0‡	5	5
		< 2	0‡	5	5
<i>Serratia marcescens</i>	43862™	95	5	5	5
		2	5	5	5
<i>Staphylococcus aureus</i>	27217™	125	5	5	5
		4	5	5	5
<i>Staphylococcus epidermidis</i>	49134™	35	5	5	5
		3	5	5	5
<i>Streptococcus viridans</i>	Clinical Isolate	43	5	5	5
		3	5	5	5
Positive			80	90	89
Total % Recovery			88.9%	100%	98.9%
95% Confidence Interval			80.5-94.5	96.0-100.0	94.0-99.9
% Recovery of Facultative Organisms and Strict Aerobes			100%	-	-
95% Confidence Interval			95.5-100.0	-	-
% Recovery of Facultative Organisms and Strict Anaerobes			-	100%	-
95% Confidence Interval			-	96.0-100.0	-

\* Prior to inoculation, 5 samples from each WBPC platelet bag were inoculated (10 mL into each bottle) into plastic BACT/ALERT® BPA and plastic BACT/ALERT® BPN culture bottles (a total of 180 samples) to verify sterility of the platelet bags (negative controls). All negative controls were negative (sterile), i.e., there were no false positives.

† Five replicates of each bottle type were inoculated with each organism at each inoculum level.



‡ High oxygen levels in this bottle prevent the growth of this organism. Recovery occurred in the BACT/ALERT® BPN bottles. **Both BACT/ALERT® BPA and BACT/ALERT® BPN culture bottles should be inoculated from the platelet specimen for optimal recovery of contaminating microorganisms.**

#### Detection of Organisms in Leukocyte-Reduced 6 Unit Pool of Whole Blood Platelet Concentrates (WBPC)

A study to determine the ability of the culture bottles to detect the presence of microorganisms in a pool of six (6) units of leukocyte-reduced pooled whole blood platelet concentrates was performed at two clinical sites. Platelet bags were seeded at Day 2 with 10 individual microorganisms to include:

- *Bacillus cereus* ATCC® 11778™
- *Escherichia coli* ATCC® 25922™
- *Enterobacter cloacae* clinical isolate
- *Klebsiella pneumoniae* clinical isolate
- *Cutibacterium acnes* ATCC® 11827™
- *Serratia marcescens* ATCC® 43862™
- *Staphylococcus aureus* ATCC® 27217™
- *Staphylococcus epidermidis* ATCC® 49134™
- *Streptococcus viridans* group clinical isolate
- *Clostridium perfringens* ATCC® 13124™

Ten replicates of each bottle type at each site were inoculated (4 mL) with each organism at each inoculum level. Each organism was seeded into one platelet unit at a target level of 10 and 100 CFU/mL, and then that unit was pooled with five sterile units. The concentration of organisms in the pool varied for each organism and ranged from <2 to 33 CFU/mL. An additional 207 BACT/ALERT® BPA and 207 BACT/ALERT® BPN bottles served as negative controls (LRWBPC with no seeded organisms added). One false positive was detected at Site B. See the following tables for results.

**Table 3: Recovery of Organisms in a 6 Unit Pool of Leukocyte-Reduced Whole Blood Platelet Concentrates from Site A**

Microorganism	ATCC® Number	Inoculum in Pooled Unit (CFU/mL)*	Number of Positive Cultures			
			BACT/ALERT® BPA†	BACT/ALERT® BPN†	BACT/ALERT® BPA + BACT/ALERT® BPN‡	Solid Media
<i>Bacillus cereus</i>	11778™	< 2	10	10	10	6
		8	10	10	10	10
<i>Escherichia coli</i>	25922™	< 2	10	10	10	6
		6	10	10	10	10
<i>Enterobacter cloacae</i>	Clinical Isolate	2	10	10	10	10
		24	10	10	10	10
<i>Klebsiella pneumoniae</i>	Clinical Isolate	< 2	10	8	10	0
		2	10	10	10	9
<i>Cutibacterium acnes</i>	11827™	< 2	0§	10	10	2
		2.5	0§	10	10	10
<i>Serratia marcescens</i>	43862™	< 2	4	4	7	1
		< 2	10	10	10	6
<i>Staphylococcus aureus</i>	27217™	2	10	10	10	4
		10	10	10	10	10
<i>Staphylococcus epidermidis</i>	49134™	2	10	10	10	6
		11	10	10	10	10
<i>Streptococcus viridans</i>	Clinical Isolate	< 2	9	9	10	1
		< 2	10	10	10	10

Microorganism	ATCC® Number	Inoculum in Pooled Unit (CFU/mL)*	Number of Positive Cultures			
			BACT/ALERT® BPA†	BACT/ALERT® BPN†	BACT/ALERT® BPA + BACT/ALERT® BPN‡	Solid Media
<i>Clostridium perfringens</i>	13124™	< 2	0§	3	3	2
		< 2	0§	10	10	1
Positive			153	184	190	124
Total % Recovery			76.5%	92.0%	95.0%	62.0%
95% Confidence Interval			70.0-82.2	87.3-95.4	91.0-97.6	54.9-68.8
% Recovery of Facultative Organisms and Strict Aerobes			95.6%	-	-	-
95% Confidence Interval			91.2-98.2	-	-	-
% Recovery of Facultative Organisms and Strict Anaerobes			-	92.0%	-	-
95% Confidence Interval			-	87.3-95.4	-	-

\* One hundred and five bottles each of BACT/ALERT® BPA and BACT/ALERT® BPN were inoculated with 10 mL of non-seeded, sterile, leukocyte-reduced WBPC and incubated along with the seeded, inoculated bottles to serve as negative controls. These additional negative controls were also utilized to establish that leukocyte-reduced WBPC at a 10 mL sample volume do not cause a high rate of false positive results to occur. These negative controls (6 unit pool) also represent the upper sample volume range of leukocyte-reduced WBPC that can be inoculated into the BACT/ALERT® bottles. All negative controls were negative (sterile), i.e., there were no false positives at this site.

† Ten replicates of each bottle type were inoculated with each organism at each inoculum level.

‡ Number of positive results when one or both of the paired BACT/ALERT® bottles (BACT/ALERT® BPA and BACT/ALERT® BPN) were detected. For best overall recovery when culturing platelet specimens it is strongly recommended that more than one type of culture bottle be utilized (e.g., one aerobic and one anaerobic).

§ High oxygen levels in this bottle prevent the growth of these organisms. Recovery occurred in the BACT/ALERT® BPN bottles. **Both BACT/ALERT® BPA and BACT/ALERT® BPN culture bottles should be inoculated from the platelet specimen for optimal recovery of contaminating microorganisms.**

**Table 4: Recovery of Organisms in a 6 Unit Pool of Leukocyte-Reduced Whole Blood Platelet Concentrates from Site B**

Microorganism	ATCC® Number	Inoculum in Pooled Unit (CFU/mL)*	Number of Positive Cultures			
			BACT/ALERT® BPA†	BACT/ALERT® BPN†	BACT/ALERT® BPA + BACT/ALERT® BPN‡	Solid Media
<i>Bacillus cereus</i>	11778™	3	10	10	10	6
		11	10	10	10	10
<i>Escherichia coli</i>	25922™	3	10	10	10	6
		10	10	10	10	10
<i>Enterobacter cloacae</i>	Clinical Isolate	2	10	10	10	10
		14	10	10	10	10
<i>Klebsiella pneumoniae</i>	Clinical Isolate	< 2	5	6	7	0
		3	10	10	10	3
<i>Cutibacterium acnes</i>	11827™	< 2	0§	0	0	0
		18	1	10	10	10
<i>Serratia marcescens</i>	43862™	4	10	10	10	8
		31	10	10	10	10

Microorganism	ATCC® Number	Inoculum in Pooled Unit (CFU/mL)*	Number of Positive Cultures			
			BACT/ALERT® BPA†	BACT/ALERT® BPN†	BACT/ALERT® BPA + BACT/ALERT® BPN‡	Solid Media
<i>Staphylococcus aureus</i>	27217™	< 2	10	10	10	9
		17	10	10	10	10
<i>Staphylococcus epidermidis</i>	49134™	2	10	10	10	8
		19	10	10	10	10
<i>Streptococcus viridans</i>	Clinical Isolate	1	10	10	10	10
		33	10	10	10	10
<i>Clostridium perfringens</i>	13124™	< 2	0§	2	2	8
		15	0§	10	10	10
Positive			156	178	179	162
Total % Recovery			78.0%	89.0%	89.5%	81.0%
95% Confidence Interval			71.6-83.5	83.8-93.0	84.4-93.4	74.9-86.2
% Recovery of Facultative Organisms and Strict Aerobes			96.9%	-	-	-
95% Confidence Interval			92.9-99.0	-	-	-
% Recovery of Facultative Organisms and Strict Anaerobes			-	89.0%	-	-
95% Confidence Interval			-	83.8-93.0	-	-

\* One hundred and two bottles each of BACT/ALERT® BPA and BACT/ALERT® BPN were inoculated with 10 mL of non-seeded, sterile, leukocyte-reduced WBPC and incubated along with the seeded, inoculated bottles to serve as negative controls. These additional negative controls were also utilized to establish that leukocyte-reduced WBPC at a 10 mL sample volume do not cause a high rate of false positive results to occur. These negative controls (6 unit pool) also represent the upper sample volume range of leukocyte-reduced WBPC that can be inoculated into the BACT/ALERT® bottles. One negative control was positive and was determined to be a true false positive at this site.

† Ten replicates of each bottle type were inoculated with each organism at each inoculum level.

‡ Number of positive results when one or both of the paired BACT/ALERT® bottles (BACT/ALERT® BPA and BACT/ALERT® BPN) were detected. For best overall recovery when culturing platelet specimens it is strongly recommended that more than one type of culture bottle be utilized (e.g., one aerobic and one anaerobic).

§ High oxygen levels in this bottle prevent the growth of these organisms. Recovery occurred in the BACT/ALERT® BPN bottles. **Both BACT/ALERT® BPA and BACT/ALERT® BPN culture bottles should be inoculated from the platelet specimen for optimal recovery of contaminating microorganisms.**

#### Detection of Organisms in Pooled Leukocyte-Reduced Whole Blood Derived Buffy Coat Platelet Concentrates

A study to determine the ability of the culture bottles to detect the presence of microorganisms in leukocyte-reduced pooled whole blood derived buffy coat platelet concentrates was performed. Pooled buffy coat platelet concentrates were prepared either in plasma only or with PAS (Platelet Additive Solution- Composol PS Fresenius Kabi, Bad Homburg, Germany) added, comprising 65% PAS and 35% plasma. Each platelet unit in the trial was prepared from a homogeneous multi pool of platelet concentrates (n=4 normal pools) in order to reduce variability from plasma. Bags were seeded at Day 2 with 10 individual microorganisms to include:

- *Bacillus cereus* ATCC® 7064™
- *Enterobacter cloacae* ATCC® 29005™
- *Escherichia coli* ATCC® 25922™
- *Klebsiella pneumoniae* ATCC® 8045™
- *Pseudomonas aeruginosa* ATCC® 27853™
- *Salmonella enterica* subsp. *enterica* serovar Heidelberg ATCC® 8326™
- *Serratia marcescens* ATCC® 43862™
- *Staphylococcus aureus* ATCC® 27217™
- *Staphylococcus epidermidis* ATCC® 49134™

- *Streptococcus agalactiae* ATCC® 12927™

The pool was sampled, via 7-10 mL inocula, into each bottle of a BACT/ALERT BPA® and BACT/ALERT® BPN bottle set. Six (6) such sets or “repetitions” were performed for each of the 10 CFU/mL target seeded pools. Each seeded pool was also inoculated onto Blood Agar Plates (BAPs), which acted as both the predicate device, and as a colony count plate. An additional 8 bottles served as negative controls (buffy coat platelets with no seeded organisms added). No false positives or contaminated negative controls were detected. See the following tables for results.

**Table 5: Recovery of Microorganisms in Pooled Leukocyte-Reduced Whole Blood Derived Buffy Coat Platelet Concentrates with BACT/ALERT® BPA and BACT/ALERT® BPN – Plasma Only Method**

Microorganism	ATCC® Number	Inoculum in Pooled Unit (CFU/mL)†	Number of Positive Cultures			
			BACT/ALERT® BPA†	BACT/ALERT® BPN†	BACT/ALERT® BPA + BACT/ALERT® BPN Set‡	Plate Media
<i>Bacillus cereus</i>	7064™	7	6	6	12	11
<i>Enterobacter cloacae</i>	29005™	23	6	6	12	12
<i>Escherichia coli</i>	25922™	13	6	6	12	12
<i>Klebsiella pneumoniae</i>	8045™	6	6	6	12	10
<i>Pseudomonas aeruginosa</i> §	27853™	16	6	0	6	12
<i>Salmonella enterica</i> subsp. <i>enterica</i> serovar Heidelberg	8326™	5	6	6	12	11
<i>Serratia marcescens</i>	43862™	14	6	6	12	12
<i>Staphylococcus aureus</i>	27217™	10	6	6	12	12
<i>Staphylococcus epidermidis</i>	49134™	5	6	6	12	12
<i>Streptococcus agalactiae</i>	12927™	15	6	6	12	12
Positive			60	54	114	116
Total % Recovery			100%	90.0%	95.0%	96.7%
95% Confidence Interval			94.0-100.0	79.5-96.2	89.4-98.1	91.7-99.1
% Recovery of Facultative Organisms			-	100% (54/54)	100% (108/108)	-
95% Confidence Interval			-	93.4-100.0	96.0-100.0	-

\* Eight bottles each of BACT/ALERT® BPA and BACT/ALERT® BPN were inoculated with 7-10 mL of non-seeded, sterile, leukocyte-reduced buffy coat Platelet Concentrates (BCPC) and incubated with the seeded, inoculated bottles to serve as negative controls. These negative controls were used to establish that leukocyte-reduced BCPC at a 7-10 mL inoculum volume do not cause false positive results to occur. All negative controls were negative (sterile), i.e., there were no false positives.

† Six replicates of each bottle type were inoculated with each organism at the 10 CFU/mL target inoculum level.

‡ Six replicates of each bottle type were inoculated with each organism at the 10 CFU/mL target inoculum level for a total of twelve (12) replicates for each organism, per two (2) bottle set, at the target inoculum level.

§ Strict Aerobe - not expected to recover in anaerobic culture bottles (BACT/ALERT® BPN).

**Table 6: Recovery of Organisms in Pooled Leukocyte-Reduced Whole Blood Derived Buffy Coat Platelet Concentrates with BACT/ALERT® BPA, BACT/ALERT® BPN, and Plate Media - Plasma/Composol\*\* Method**

Microorganism	ATCC® Number	Inoculum in Pooled Unit (CFU/mL) <sup>†</sup>	Number of Positive Cultures			
			BACT/ALERT® BPA <sup>†</sup>	BACT/ALERT® BPN <sup>†</sup>	BACT/ALERT® BPA + BACT/ALERT® BPN Set <sup>‡</sup>	Plate Media
<i>Bacillus cereus</i>	7064™	7	6	6	12	11
<i>Enterobacter cloacae</i>	29005™	23	6	6	12	12
<i>Escherichia coli</i>	25922™	13	6	6	12	12
<i>Klebsiella pneumoniae</i>	8045™	6	5	6	11	9
<i>Pseudomonas aeruginosa</i> <sup>§</sup>	27853™	16	6	1	7	12
<i>Salmonella enterica</i> subsp. <i>enterica</i> serovar Heidelberg	8326™	5	6	6	12	11
<i>Serratia marcescens</i>	43862™	14	6	6	12	12
<i>Staphylococcus aureus</i>	27217™	10	6	6	12	12
<i>Staphylococcus epidermidis</i>	49134™	5	6	6	12	12
<i>Streptococcus agalactiae</i>	12927™	15	6	6	12	12
Positive			59	55	114	114
Total % Recovery			98.3%	91.7%	95.0%	95.0%
95% Confidence Interval			91.1-99.9	81.6-97.2	89.4-98.1	89.4-98.1
% Recovery of Facultative Organisms			-	100% (54/54)	99.1% (107/108)	-
95% Confidence Interval			-	93.4-100.0	94.9-99.9	-

\*\* "Results in the table above were generated using pooled buffy coat platelets in 35% plasma and 65% platelet additive solution (Composol® PS platelet storage solution, Fresenius Kabi, Bad Homburg, Germany). Results above only represent data for the specific plasma/Composol® ratio as indicated. Other ratios of plasma and platelet additive solution or usage of platelet additive solutions with a different composition or from another manufacturer were not studied and might show different microorganism recovery results."

\* Eight bottles each of BACT/ALERT® BPA and BACT/ALERT® BPN were inoculated with 7-10 mL of non-seeded, sterile, leukocyte-reduced buffy coat Platelet Concentrates (BCPC) and incubated with the seeded, inoculated bottles to serve as negative controls. These negative controls were used to establish that leukocyte-reduced BCPC at a 7-10 mL inoculum volume do not cause false positive results to occur. All negative controls were negative (sterile), i.e., there were no false positives.

† Six replicates of each bottle type were inoculated with each organism at the 10 CFU/mL target inoculum level.

‡ Six replicates of each bottle type were inoculated with each organism at the 10 CFU/mL target inoculum level for a total of twelve (12) replicates for each organism, per two (2) bottle set, at the target inoculum level.

§ Strict Aerobe - not expected to recover in anaerobic culture bottles (BACT/ALERT® BPN).

#### Equivalency of Platelet Preparation Methods for Performance Validation Testing of the BACT/ALERT Microbial Detection Systems

A retrospective review of time to detection (TTD) data collected during performance validation testing of the BACT/ALERT® 3D and BACT/ALERT® VIRTUO® was conducted to determine if platelet preparation method had an impact on the TTD of

microorganisms seeded into leukocyte-reduced apheresis platelets (LRAP), leukocyte-reduced whole blood platelet concentrates (LRWBPC), or leukocyte-reduced whole blood derived buffy coat platelets (LRWBDBCP). Since recovery of seeded microorganisms is well established for the BACT/ALERT® Microbial Detection Systems, TTD data were used to evaluate the platelet preparation methods. The goal of this data review was to establish the equivalency of platelet preparation methods for use in seeded studies to establish performance for the BACT/ALERT® Microbial Detection Systems.

The data reviewed was collected from four independent studies conducted at two external sites where different instruments, different bottle lots, and different volumes of seeded sample (8 mL or 4 mL) were tested. All studies used both aerobic (BPA) and anaerobic (BPN) culture bottles. Site 1 tested eight microorganisms resulting in 14 combinations of microorganism-bottle-platelet concentrate (LRAP or LRWBPC) on the BACT/ALERT® 3D and the BACT/ALERT® VIRTUO®. Site 2 tested nine microorganisms resulting in 17 combinations of microorganism-bottle-platelet concentrate (LRAP or LRWBPC) on the BACT/ALERT® 3D. See the following table for results.

**Table 7: Comparison of the Average of the Mean TTD by Platelet Type on the BACT/ALERT® Systems**

Instrument   Platelet Type	LRAP TTD Hours <sup>2</sup>	LRWBDBCP / LRWBPC TTD Hours <sup>1</sup>
BACT/ALERT® VIRTUO® (Site 1 - Studies 1 & 2)	11.1	11.8
BACT/ALERT® 3D (Site 1 - Studies 1 & 2)	14.3	14.1
BACT/ALERT® 3D (Site 2 - Studies 3 & 4)	16.5	17.0 <sup>2</sup>

<sup>1</sup>8 mL Sample Volume

<sup>2</sup>4 mL Sample Volume

The results show that the Average of the Mean TTD across platelet type tested on the BACT/ALERT® 3D or BACT/ALERT® VIRTUO® indicate there are no practical differences (LRAP or LRWBDBCP / LRWBPC, <1 hour) in TTD between platelet type. The variability observed and reported in four independent studies for the average of the mean TTD across instruments and sites (LRAP ≤5.4 hours, LRWBDBCP / LRWBPC ≤5.2 hours) supports that platelet preparation method does not have a practical impact on TTD and that site-to-site and instrument-to-instrument differences have a greater impact on TTD.

Platelet preparation method does not impact the ability of the BACT/ALERT® Microbial Detection Systems to detect microorganisms; therefore, data collected during performance validation testing of the BACT/ALERT® VIRTUO using LRAP and LRWBDBCP is representative of performance of the BACT/ALERT® VIRTUO for testing LRWBPC.

### **BACT/ALERT® VIRTUO® Microbial Detection Systems**

#### **Reproducibility Results from Pooled LRWBDBCP Concentrates with BACT/ALERT® VIRTUO®**

Data in the following table represent results from seeded platelet studies conducted with two lots of BACT/ALERT® BPA and BACT/ALERT® BPN culture bottles at three sites (one external and two internal), testing a total of 270 replicates over 3 days, with two operators at each site. Reproducibility was evaluated on each of six microorganisms; five were inoculated into each bottle type (four facultative microorganisms and one obligate aerobic microorganism in BACT/ALERT® BPA and four facultative microorganisms and one obligate anaerobic microorganism in BACT/ALERT® BPN). All microorganisms were prepared from BIOBALL® products except for *P. aeruginosa* ATCC® 27853™ which was prepared as a serial dilution. The target inoculum was 5 CFU/mL of platelets with 4 mL of seeded platelets inoculated into each bottle. Percent recovery reflects a positive flag by the instrument and Gram-stain/subculture consistent with the seeded microorganism.

The following table shows the percent recovery for each of the microorganisms tested as well as the overall percent recovery by site and for all sites combined. The mean time to detection and corresponding range are shown for each microorganism for all sites, as well as the actual inoculum ranges. Leukocyte-reduced whole blood derived buffy coat platelets (LRWBDBCP) reproducibility was conducted using leukocyte-reduced apheresis platelet (LRAP) at the internal sites, as LRWBDBCP are not available in the United States.

**Table 8: Reproducibility Detection Rates of Microorganisms in Pooled Leukocyte-Reduced Whole Blood Derived Buffy Coat Platelet Concentrates**

Microorganism	% Recovery (95% Confidence Interval)				Time to Detection (hours)		Inoculum Range (CFU/mL) <sup>*</sup>
	Site 1	Site 2	Site 3	Overall	Mean	Range	
<i>Staphylococcus aureus</i> NCTC 10788	100.0% (18/18)	100.0% (18/18)	100.0% (18/18)	100.0%	13.8	12.2-15.6	2-9
				[93.4-100.0]%			
				(54/54)			
<i>Streptococcus pyogenes</i> NCTC 12696	100.0% (18/18)	100.0% (18/18)	100.0% (18/18)	100.0%	11.6	10.0-14.1	< 1-9
				[93.4-100.0]%			
				(54/54)			
<i>Escherichia coli</i> NCTC 12241	100.0% (18/18)	100.0% (18/18)	100.0% (18/18)	100.0%	9.5	8.7-10.7	2-10
				[93.4-100.0]%			
				(54/54)			
<i>Bacillus cereus</i> NCTC 7464	100.0% (18/18)	100.0% (18/18)	100.0% (18/18)	100.0%	8.8	7.0-12.3	1-8
				[93.4-100.0]%			
				(54/54)			
<i>Pseudomonas aeruginosa</i> †‡ NCTC 12924 ATCC® 27853™	100.0% (9/9)	100.0% (9/9)	100.0% (9/9)	100.0%	13.7	13.0-14.6	< 1 (Sites 1 and 2)
				[87.2-100.0]%			7-20
				(27/27)			(Site 3)
<i>Clostridium perfringens</i> NCTC 8798§	100.0% (9/9)	100.0% (9/9)	100.0% (9/9)	100.0%	8.6	7.1-11.8	2-6
				[87.2-100.0]%			
				(27/27)			
Overall	100.0%	100.0%	100.0%	100.0%	-	-	-
	[96.0-100.0]%	[96.0-100.0]%	[96.0-100.0]%	[98.6-100.0]%	-	-	-
	(90/90)	(90/90)	(90/90)	(270/270)	-	-	-

\* 54 bottles (27 BACT/ALERT® BPA and 27 BACT/ALERT® BPN) were inoculated with 10 mL of unseeded platelet to serve both as sterility checks for the platelet and as negative controls to ensure platelets inoculated into BACT/ALERT® BPA and BACT/ALERT® BPN bottles do not cause false positives to occur. Of the 27 BACT/ALERT® BPA bottles, 18 (8 into lot 1 and 10 into lot 2) were inoculated with LRAP and 9 (4 into lot 1 and 5 into lot 2) with LRWBDBCP. Of the 27 BACT/ALERT® BPN bottles, 18 (10 into lot 1 and 8 into lot 2) were inoculated with LRAP and 9 (5 into lot 1 and 4 into lot 2) with LRWBDBCP.

**Note:** LRAP and buffy coat platelets were determined to be equivalent matrices for the reproducibility seeded study.

† Site 3 substituted *P. aeruginosa* ATCC® 27853™ for *P. aeruginosa* BIOBALL® NCTC 12924 because the BIOBALL® microorganism grew inconsistently in LRWBDBCP due to plasma sensitivity.<sup>14</sup>

‡ Strict Aerobe - not tested in anaerobic culture bottles (BACT/ALERT® BPN).

§ Strict Anaerobe - not tested in aerobic culture bottles (BACT/ALERT® BPA).

**Note:** The percent recovery results, per microorganism, were 100% for all sites combined. The percent recovery result for BACT/ALERT® BPA lot 1 was 100.0% (60/60) with 95% CI: [94.0-100.0] and for BACT/ALERT® BPA lot 2 was 100.0% (75/75) with 95% CI: [95.2-100.0]%. The percent recovery result for BACT/ALERT® BPN lot 1 was 100.0% (75/75) with 95% CI: [95.2-100.0] and for BACT/ALERT® BPN lot 2 was 100.0% (60/60) with 95% CI: [94.0-100.0]%.  
**Note:** Overall, the negative agreement rate for BACT/ALERT® VIRTUO® was 100% (54/54). The negative agreement rate for BACT/ALERT® BPA lot 1 was 100%, (12/12) of the negative controls were declared negative by BACT/ALERT® VIRTUO®.

<sup>14</sup> Adamik M, Anheuser MB, Klein J, Deol P. Determination of Plasma Resistance of Microorganisms Used for BacT/ALERT® VIRTUO® Detection System Platelet Study. Poster presented at: American Association of Blood Banks (AABB) Annual Meeting 2017; October, 2017; San Diego, CA.

The negative agreement rate for BACT/ALERT® BPA lot 2 was 100%, (15/15) of the negative controls were declared negative by BACT/ALERT® VIRTUO®. The negative agreement rate for BACT/ALERT® BPN lot 1 was 100%, (15/15) of the negative controls were declared negative by BACT/ALERT® VIRTUO®. The negative agreement rate for BACT/ALERT® BPN lot 2 was 100%, (12/12) of the negative controls were declared negative by BACT/ALERT® VIRTUO®.

### Reproducibility Results from LRAP with BACT/ALERT® VIRTUO®

Data in the following table represent results from seeded leukocyte-reduced apheresis platelet (LRAP) units conducted with two lots of BACT/ALERT® BPA and BACT/ALERT® BPN culture bottles at three sites (two external and one internal), testing a total of 270 replicates over 3 days, with two operators at each site. Reproducibility was evaluated on each of six microorganisms; five were inoculated into each bottle type (four facultative microorganisms and one obligate aerobic microorganism in BACT/ALERT® BPA and four facultative microorganisms and one obligate anaerobic microorganism in BACT/ALERT® BPN). All microorganisms were prepared from BIOBALL® products except for *P. aeruginosa* ATCC® 27853™ which was prepared as a serial dilution. The target inoculum was 5 CFU/mL of platelets with 4 mL of seeded platelets inoculated into each bottle. Percent recovery reflects a positive flag by the instrument and Gram-stain/subculture consistent with the seeded microorganism and from each bottle set identification was performed.

The following table shows the percent recovery for each of the microorganisms tested as well as the overall percent recovery by site and for all sites combined. The mean time to detection and corresponding range are shown for each microorganism for all sites, as well as the actual inoculum ranges.

**Table 9: Reproducibility Detection Rates of Microorganisms in Leukocyte-Reduced Apheresis Platelets**

Microorganism	% Recovery (95% Confidence Interval)				Time to Detection (hours)		Inoculum Range (CFU/mL)*
	Site 1	Site 2	Site 3	Overall	Mean	Range	
<i>Staphylococcus aureus</i> NCTC 10788	100% (18/18)	100% (18/18)	100% (18/18)	100%	13.2	9.6-15.9	1-7
				[93.4-100]%			
				(54/54)			
<i>Streptococcus pyogenes</i> NCTC 12696	100% (18/18)	100% (18/18)	100% (18/18)	100%	11.7	9.8-16.9	1-7
				[93.4-100]%			
				(54/54)			
<i>Escherichia coli</i> NCTC 12241	100% (18/18)	100% (18/18)	100% (18/18)	100%	9.3	7.8-10.3	1-7
				[93.4-100]%			
				(54/54)			
<i>Bacillus cereus</i> NCTC 7464	100% (18/18)	100% (18/18)	100% (18/18)	100%	8.5	7.2-13.3	1-7
				[93.4-100]%			
				(54/54)			
<i>Pseudomonas aeruginosa</i> † ATCC® 27853™	100% (9/9)	100% (9/9)	100% (9/9)	100%	13.2	12.5-14.2	1-20
				[87.2-100]%			
				(27/27)			
<i>Clostridium perfringens</i> ‡ NCTC 8798	100% (9/9)	100% (9/9)	100% (9/9)	100%	7.9	6.6-9.8	1-9
				[87.2-100]%			
				(27/27)			
Overall	100%	100%	100%	100%	-	-	-
	[96.0-100]%	[96.0-100]%	[96.0-100]%	[98.6-100]%	-	-	-
	(90/90)	(90/90)	(90/90)	(270/270)	-	-	-

\* 54 bottles (27 BACT/ALERT® BPA and 27 BACT/ALERT® BPN) were inoculated with 10 mL of unseeded LRAP to serve both as sterility checks for the platelets and as negative controls to ensure platelets inoculated into BACT/ALERT® BPA and BACT/ALERT® BPN bottles do not cause false positives to occur. Of the 27 BACT/ALERT® BPA bottles, 12 were from lot 1 and 15 were from lot 2. Of the 27 BACT/ALERT® BPN bottles, 15 were from lot 1 and 12 were from lot 2.

† Strict Aerobe - not tested in anaerobic culture bottles (BACT/ALERT® BPN).



‡ Strict Anaerobe - not tested in aerobic culture bottles (BACT/ALERT® BPA).

**Note:** The percent recovery results, per microorganism, were 100% for all sites combined.

**Note:** The percent recovery result for BACT/ALERT® BPA lot 1 was 100% (60/60) with 95% CI: [94.0-100]% and for BACT/ALERT® BPA lot 2 was 100% (75/75) with 95% CI: [95.2-100]%.

**Note:** The percent recovery result for BACT/ALERT® BPN lot 1 was 100% (75/75) with 95% CI: [95.2-100]% and for BACT/ALERT® BPN lot 2 was 100% (60/60) with 95% CI: [94.0-100]%.

**Note:** Overall, the negative agreement rate for BACT/ALERT® VIRTUO® was 100% (54/54). The negative agreement rate for BACT/ALERT® BPA lot 1 was 100%, (12/12) of the negative controls were declared negative by BACT/ALERT® VIRTUO®. The negative agreement rate for BACT/ALERT® BPA lot 2 was 100%, (15/15) of the negative controls were declared negative by BACT/ALERT® VIRTUO®. The negative agreement rate for BACT/ALERT® BPN lot 1 was 100%, (15/15) of the negative controls were declared negative by BACT/ALERT® VIRTUO®. The negative agreement rate for BACT/ALERT® BPN lot 2 was 100%, (12/12) of the negative controls were declared negative by BACT/ALERT® VIRTUO®.

### Clinical Study Results from Pooled LRWBDBCP Concentrates with BACT/ALERT® VIRTUO®

An external study to determine the ability of the culture bottles to detect the presence of microorganisms in pooled LRWBDBCP concentrates was performed at one (1) clinical site. The platelet concentrates were prepared either in plasma only or with PAS (Platelet Additive Solution-SSP+ (Macopharma, Mouvaux, France)) added, comprising 65% PAS and 35% plasma. Each platelet unit in the trial was prepared from a pool of up to four platelet concentrates in order to reduce variability from plasma. Aliquots of the pools were seeded with BIOBALL® microorganisms at a target inoculum of 3 CFU/mL or with microorganisms prepared from serial dilutions at a target inoculum of 10 CFU/mL. The aliquots were from Day 2 platelets and seeded with 11 individual microorganisms:

- *Bacillus cereus* NCTC 7464 Product 56023 BIOBALL® SINGLESHOT
- *Escherichia coli* NCTC 12241 Product 56035 BIOBALL® SINGLESHOT
- *Staphylococcus epidermidis* NCTC 6513 Product 56092 BIOBALL® SINGLESHOT
- *Clostridium perfringens* NCTC 8798 Product 56028 BIOBALL® SINGLESHOT
- *Staphylococcus aureus* NCTC 10788 Product 56045 BIOBALL® SINGLESHOT
- *Pseudomonas aeruginosa* ATCC® 27853™
- *Streptococcus agalactiae* ATCC® 12927™
- *Serratia liquefaciens* ATCC® 35551™
- *Enterobacter cloacae* ATCC® 29005™
- *Klebsiella pneumoniae* ATCC® 8045™
- *Salmonella enterica* subsp. *enterica* serovar Heidelberg ATCC® 8326™

Eight (8) mL of the seeded platelet aliquots were inoculated into the appropriate BACT/ALERT® BPA and BACT/ALERT® BPN bottles and the bottles were tested in the BACT/ALERT® VIRTUO® and BACT/ALERT® 3D systems. A total of 10 repetitions/system were performed for each microorganism/bottle type/platelet preparation type (i.e., plasma only, plasma with PAS) and system (BACT/ALERT® VIRTUO®, BACT/ALERT® 3D). Recovery and time to detection for test microorganisms by bottle type, platelet type, and instrument are presented in the following tables, as well as the actual inoculum ranges.

**Table 10: Recovery of Microorganisms in Pooled Leukocyte-Reduced Whole Blood Derived Buffy Coat Platelet Concentrates with BACT/ALERT® BPA and BACT/ALERT® BPN – Plasma Only Method**

Microorganism	BACT/ALERT® Culture Bottle *	Inoculum Range (CFU/mL)†	Number of Positive Cultures		Time to Detection (hours)			
			BACT/ALERT® 3D	BACT/ALERT® VIRTUO®	BACT/ALERT® 3D		BACT/ALERT® VIRTUO®	
					Mean‡	Range	Mean‡	Range
<i>Bacillus cereus</i> NCTC 7464	BACT/ALERT® BPA	2-6	10	10	10.1	9.9-10.3	8.1	7.9-8.3
	BACT/ALERT® BPN	0-4	10	10	10.7	10.1-11.6	8.5	7.9-9.0

Microorganism	BACT/ALERT® Culture Bottle* †	Inoculum Range (CFU/mL)†	Number of Positive Cultures		Time to Detection (hours)			
			BACT/ALERT® 3D	BACT/ALERT® VIRTUO®	BACT/ALERT® 3D		BACT/ALERT® VIRTUO®	
					Mean‡	Range	Mean‡	Range
<i>Clostridium perfringens</i> § NCTC 8798	BACT/ALERT® BPA	-	-	-	-	-	-	-
	BACT/ALERT® BPN	1-3	10	10	11.4	10.4-13.9	8.9	8.0-10.6
<i>Enterobacter cloacae</i> ATCC® 29005™	BACT/ALERT® BPA	20-28	10	10	14.8	14.4-15.2	12.6	12.2-13.3
	BACT/ALERT® BPN	18-29	10	10	11.4	11.2-11.4	8.9	8.6-9.1
<i>Escherichia coli</i> NCTC 12241	BACT/ALERT® BPA	2-5	10	10	12.7	12.4-13.1	10.3	10.1-10.5
	BACT/ALERT® BPN	0-5	10	10	11.6	11.4-11.7	9.9	9.5-10.1
<i>Klebsiella pneumoniae</i> ATCC® 8045™	BACT/ALERT® BPA	6-12	10	10	15.3	15.0-16.2	13.7	12.9-15.2
	BACT/ALERT® BPN	4-9	10	10	12.9	12.5-13.0	11.1	10.9-11.4
<i>Pseudomonas aeruginosa</i> ¶ ATCC® 27853™	BACT/ALERT® BPA	11-18	10	10	16.7	16.3-16.9	13.8	13.6-14.0
	BACT/ALERT® BPN	-	-	-	-	-	-	-
<i>Salmonella enterica</i> subsp. <i>enterica</i> serovar Heidelberg ATCC® 8326™	BACT/ALERT® BPA	9-18	10	10	15.6	14.9-16.1	12.6	12.3-13.4
	BACT/ALERT® BPN	7-20	10	10	10.9	10.8-11.3	8.7	8.6-8.9
<i>Serratia liquefaciens</i> ATCC® 35551™	BACT/ALERT® BPA	0-2	10	10	16.1	16.0-16.3	14.0	13.6-14.4
	BACT/ALERT® BPN	0-2	10	10	14.6	14.3-15.2	13.4	12.5-14.3
<i>Staphylococcus aureus</i> NCTC 10788	BACT/ALERT® BPA	1-3	10	10	18.0	17.7-18.4	15.2	15.0-16.0
	BACT/ALERT® BPN	2-5	10	10	15.7	15.0-16.5	13.9	12.9-14.6
<i>Staphylococcus epidermidis</i> NCTC 6513	BACT/ALERT® BPA	1-5	10	10	17.5	17.2-17.7	15.3	15.0-15.5
	BACT/ALERT® BPN	3-4	10	10	18.9	16.4-20.9	16.1	15.0-17.8
<i>Streptococcus agalactiae</i> ATCC® 12927™	BACT/ALERT® BPA	17-22	10	10	10.9	10.7-11.1	8.3	8.0-8.5
	BACT/ALERT® BPN	17-31	10	10	10.2	10.1-10.4	7.9	7.8-8.2
Positive	BACT/ALERT® BPA	-	100	100	-	-	-	-
	BACT/ALERT® BPN	-	100	100	-	-	-	-

Microorganism	BACT/ALERT® Culture Bottle *	Inoculum Range (CFU/mL)†	Number of Positive Cultures		Time to Detection (hours)			
			BACT/ALERT® 3D	BACT/ALERT® VIRTUO®	BACT/ALERT® 3D		BACT/ALERT® VIRTUO®	
					Mean‡	Range	Mean‡	Range
Total % Recovery	BACT/ALERT® BPA	-	100%	100%	-	-	-	-
	BACT/ALERT® BPN	-	100%	100%	-	-	-	-
95% Confidence Interval	BACT/ALERT® BPA	-	96.4%-100%	96.4%-100%	-	-	-	-
	BACT/ALERT® BPN	-	96.4%-100%	96.4%-100%	-	-	-	-

\* Ten replicates each of BACT/ALERT® BPA and BACT/ALERT® BPN bottles were inoculated with each microorganism at the inoculum level indicated.

† One BACT/ALERT® BPA and one BACT/ALERT® BPN bottle for each system was inoculated with 10 mL of “unseeded” buffy coat platelet product type being tested to serve as negative controls each time seeded testing was performed. These negative controls served as a sterility test of the platelet units and to establish that LRWBDBCP concentrates at a 10 mL inoculum volume do not cause false positive results to occur. The number of negative controls tested was supplemented from leftover platelets from the seeded testing and/or with additional platelet preparation such that 100 repetitions of negative controls (50 BACT/ALERT® BPA and 50 BACT/ALERT® BPN) were performed on each system (BACT/ALERT® VIRTUO® and BACT/ALERT® 3D). All negative controls were negative (sterile), i.e., there were no false positives.

‡ Mean of 10 replicates

§ Strict Anaerobe - not tested in aerobic culture bottles (BACT/ALERT® BPA).

¶ Strict Aerobe - not tested in anaerobic culture bottles (BACT/ALERT® BPN).

**Table 11: Recovery of Microorganisms in Pooled Leukocyte-Reduced Whole Blood Derived Buffy Coat Platelet Concentrates with BACT/ALERT® BPA and BACT/ALERT® BPN – Plasma/SSP+ Method**

Microorganism	BACT/ALERT® Culture Bottle *	Inoculum Range (CFU/mL)†	Number of Positive Cultures		Time to Detection (hours)			
			BACT/ALERT® 3D	BACT/ALERT® VIRTUO®	BACT/ALERT® 3D		BACT/ALERT® VIRTUO®	
					Mean‡	Range	Mean‡	Range
<i>Bacillus cereus</i> NCTC 7464	BACT/ALERT® BPA	1-5	10	10	9.8	9.5-10.0	8.0	7.8-8.0
	BACT/ALERT® BPN	1-3	10	10	11.7	10.7-12.7	9.9	8.6-10.9
<i>Clostridium perfringens</i> § NCTC 8798	BACT/ALERT® BPA	-	-	-	-	-	-	-
	BACT/ALERT® BPN	1-4	10	10	10.3	9.7-11.2	8.4	8.3-8.8
<i>Enterobacter cloacae</i> ™ ATCC® 29005	BACT/ALERT® BPA	14-28	10	10	12.3	12.2-12.6	9.8	9.7-9.9
	BACT/ALERT® BPN	13-20	10	10	11.4	11.2-11.6	9.0	8.8-9.2

Microorganism	BACT/ALERT® Culture Bottle* †	Inoculum Range (CFU/mL)†	Number of Positive Cultures		Time to Detection (hours)			
			BACT/ALERT® 3D	BACT/ALERT® VIRTUO®	BACT/ALERT® 3D		BACT/ALERT® VIRTUO®	
					Mean‡	Range	Mean‡	Range
<i>Escherichia coli</i> NCTC 12241	BACT/ALERT® BPA	1-2	10	10	12.1	11.9-12.3	10.2	9.9-10.6
	BACT/ALERT® BPN	1-4	10	10	11.1	10.9-11.3	9.6	9.3-9.8
<i>Klebsiella pneumoniae</i> ATCC® 8045™	BACT/ALERT® BPA	7-12	10	10	13.5	13.5-13.7	11.6	11.3-11.8
	BACT/ALERT® BPN	6-11	10	10	13.0	12.7-13.2	11.1	10.8-11.4
<i>Pseudomonas aeruginosa</i> †† ATCC® 27853™	BACT/ALERT® BPA	15-24	10	10	16.0	15.7-16.2	13.7	13.2-14.4
	BACT/ALERT® BPN	-	-	-	-	-	-	-
<i>Salmonella enterica</i> subsp. <i>enterica</i> serovar Heidelberg ATCC® 8326™	BACT/ALERT® BPA	21-31	10	10	12.0	11.7-12.2	9.8	9.6-10.1
	BACT/ALERT® BPN	15-23	10	10	10.9	10.7-11.0	9.0	8.7-9.2
<i>Serratia liquefaciens</i> ATCC® 35551™	BACT/ALERT® BPA	0-2	10	10	16.3	16.1-16.6	14.2	13.8-14.7
	BACT/ALERT® BPN	1-3	10	10	16.8	16.0-17.3	15.1	14.5-15.5
<i>Staphylococcus aureus</i> NCTC 10788	BACT/ALERT® BPA	2-7	10	10	17.3	17.1-17.4	14.8	14.5-15.5
	BACT/ALERT® BPN	2-6	10	10	16.2	15.6-16.7	14.2	13.6-15.1
<i>Staphylococcus epidermidis</i> NCTC 6513	BACT/ALERT® BPA	3-6	10	10	17.9	17.6-18.1	16.0	15.7-16.5
	BACT/ALERT® BPN	1-6	10	10	18.9	17.8-19.8	18.8	14.7-19.7
<i>Streptococcus agalactiae</i> ATCC® 12927™	BACT/ALERT® BPA	13-24	10	10	10.4	10.3-10.4	8.6	8.5-8.7
	BACT/ALERT® BPN	17-21	10	10	10.2	9.9-10.4	8.3	8.1-8.6
Positive	BACT/ALERT® BPA	-	100	100	-	-	-	-
	BACT/ALERT® BPN	-	100	100	-	-	-	-
Total % Recovery	BACT/ALERT® BPA	-	100%	100%	-	-	-	-
	BACT/ALERT® BPN	-	100%	100%	-	-	-	-
95% Confidence Interval	BACT/ALERT® BPA	-	96.4%-100%	96.4%-100%	-	-	-	-
	BACT/ALERT® BPN	-	96.4%-100%	96.4%-100%	-	-	-	-

\* Ten replicates each of BACT/ALERT® BPA and BACT/ALERT® BPN bottles were inoculated with each microorganism at the inoculum level indicated.

† One BACT/ALERT® BPA and one BACT/ALERT® BPN bottle for each system was inoculated with 10 mL of “unseeded” buffy coat platelet product type being tested to serve as negative controls each time seeded testing was performed. These negative controls served as a sterility test of the platelet units and to establish that LRWBDBCP concentrates at a 10 mL inoculum volume do not cause false positive results to occur. The number of negative controls tested was supplemented from leftover platelets from the seeded testing and/or with additional platelet preparation such that 111 repetitions of negative controls (56 BACT/ALERT® BPA and 55 BACT/ALERT® BPN) were performed on BACT/ALERT® 3D and 110 repetitions of negative controls (55 BACT/ALERT® BPA and 55 BACT/ALERT® BPN) were performed on BACT/ALERT® VIRTUO®. One (1) BACT/ALERT® BPN negative control was signaled positive by BACT/ALERT® 3D and was determined to be a true false positive (bottle subculture negative).

‡ Mean of 10 replicates

§ Strict Anaerobe - not tested in aerobic culture bottles (BACT/ALERT® BPA).

¶ Strict Aerobe - not tested in anaerobic culture bottles (BACT/ALERT® BPN).

#### Clinical Study Results from LRAP with BACT/ALERT® VIRTUO®

A study to determine the ability of the BACT/ALERT® BPA and BACT/ALERT® BPN culture bottles to detect the presence of microorganisms in LRAP in plasma only was performed at two (2) external sites. The platelets were obtained from inventory of a platelet collection agency. Aliquots of the platelets were seeded with BIOBALL® microorganisms at a target inoculum of 5 CFU/mL or with microorganisms prepared from serial dilutions at a target inoculum of 10 CFU/mL. The aliquots from platelets were seeded with 11 individual microorganisms:

- *Bacillus cereus* NCTC 7464 Product 56023 BIOBALL® SINGLESHOT
- *Clostridium perfringens* NCTC 8798 Product 56028 BIOBALL® SINGLESHOT
- *Escherichia coli* NCTC 12241 Product 56035 BIOBALL® SINGLESHOT
- *Staphylococcus epidermidis* NCTC 6513 Product 56092 BIOBALL® SINGLESHOT
- *Staphylococcus aureus* NCTC 10788 Product 56045 BIOBALL® SINGLESHOT
- *Streptococcus pyogenes* NCTC 12696 Product 56046 BIOBALL® SINGLESHOT
- *Enterobacter cloacae* ATCC® 29005™
- *Klebsiella pneumoniae* ATCC® 8045™
- *Serratia marcescens* ATCC® 43862™
- *Streptococcus sanguinis* ATCC® 10556™
- *Pseudomonas aeruginosa* ATCC® 27853™

Four (4) mL of the seeded LRAP aliquots were inoculated into the appropriate BACT/ALERT® BPA and BACT/ALERT® BPN bottles and the bottles were tested in the BACT/ALERT® VIRTUO® and BACT/ALERT® 3D systems. A total of 10 repetitions/system were performed for each microorganism/bottle type/platelet preparation type (LRAP) and system ( BACT/ALERT® VIRTUO®, BACT/ALERT® 3D). Recovery and time to detection for test microorganisms by bottle type, platelet type, and instrument are presented in the following table, as well as the actual inoculum ranges.

**Table 12: Recovery of Microorganisms in Leukocyte-Reduced Apheresis Platelets with BACT/ALERT® BPA and BACT/ALERT® BPN – Sites Combined**

Microorganism	BACT/ALERT® Culture Bottle*	Inoculum Ranges (CFU/mL)†	Number of Positive Cultures		Time to Detection (hours)			
			BACT/ALERT® 3D	BACT/ALERT® VIRTUO®	BACT/ALERT® 3D		BACT/ALERT® VIRTUO®	
					Mean‡	Range	Mean‡	Range
<i>Bacillus cereus</i> NCTC 7464	BACT/ALERT® BPA	1-7	20	19	10.3	10.1-10.7	7.7§	7.3-8.2
	BACT/ALERT® BPN	1-8	20	20	11.8	9.8-16.7	9.4	7.8-12.0

Microorganism	BACT/ALERT® Culture Bottle* †	Inoculum Ranges (CFU/mL)†	Number of Positive Cultures		Time to Detection (hours)			
			BACT/ALERT® 3D	BACT/ALERT® VIRTUO®	BACT/ALERT® 3D		BACT/ALERT® VIRTUO®	
					Mean‡	Range	Mean‡	Range
<i>Clostridium perfringens</i> ¶ NCTC 8798	BACT/ALERT® BPA	-	-	-	-	-	-	-
	BACT/ALERT® BPN	1-7	20	20	11.1	10.7-12.2	8.0	7.1-9.1
<i>Enterobacter cloacae</i> ATCC® 29005™	BACT/ALERT® BPA	3-16	20	20	12.7	12.5-13.2	10.0	9.8-10.8
	BACT/ALERT® BPN	4-20	20	20	11.6	11.3-12.0	9.0	8.8-9.3
<i>Escherichia coli</i> NCTC 12241	BACT/ALERT® BPA	1-8	20	20	12.6	12.0-12.9	9.6	9.1-10.2
	BACT/ALERT® BPN	1-7	20	20	11.9	11.5-12.5	9.2	8.6-9.8
<i>Klebsiella pneumoniae</i> ATCC® 8045™	BACT/ALERT® BPA	2-17	20	20	14.5	13.9-15.4	11.7	11.0-12.6
	BACT/ALERT® BPN	3-19	20	20	13.8	13.0-14.4	11.2	10.6-12.1
<i>Pseudomonas aeruginosa</i> ** ATCC® 27853™	BACT/ALERT® BPA	2-20	20	20	16.7	16.3-17.0	13.3	12.7-13.9
	BACT/ALERT® BPN	-	-	-	-	-	-	-
<i>Serratia marcescens</i> ATCC® 43862™	BACT/ALERT® BPA	2-14	20	20	12.4	11.7-13.9	10.2	9.7-10.9
	BACT/ALERT® BPN	1-12	20	20	12.5	11.8-13.0	10.2	9.3-10.8
<i>Staphylococcus aureus</i> NCTC 10788	BACT/ALERT® BPA	1-10	20	20	17.1	16.6-17.3	14.5	13.7-15.9
	BACT/ALERT® BPN	2-6	20	20	15.7	14.6-17.1	12.5	11.8-13.3
<i>Staphylococcus epidermidis</i> NCTC 6513	BACT/ALERT® BPA	1-7	20	20	20.4	19.6-21.4	18.1	17.2-18.9
	BACT/ALERT® BPN	1-8	20	20	17.7	16.3-20.8	14.6	13.2-15.9
<i>Streptococcus pyogenes</i> NCTC 12696	BACT/ALERT® BPA	1-7	20	20	15.0	14.2-15.8	12.2	11.4-13.4
	BACT/ALERT® BPN	2-9	20	20	13.9	13.4-14.4	11.2	10.5-12.0
<i>Streptococcus sanguinis</i> ATCC® 10556™	BACT/ALERT® BPA	1-16	20	20	21.8	19.2-26.4	17.4	13.8-22.0
	BACT/ALERT® BPN	1-17	20	19	27.0	19.2-39.6	20.7§	13.2-34.3
Positive	BACT/ALERT® BPA	-	200	199	-	-	-	-
	BACT/ALERT® BPN	-	200	199	-	-	-	-

Microorganism	BACT/ALERT® Culture Bottle*	Inoculum Ranges (CFU/mL)†	Number of Positive Cultures		Time to Detection (hours)			
			BACT/ALERT® 3D	BACT/ALERT® VIRTUO®	BACT/ALERT® 3D		BACT/ALERT® VIRTUO®	
					Mean‡	Range	Mean‡	Range
Total % Recovery	BACT/ALERT® BPA	-	100%	99.5%	-	-	-	-
	BACT/ALERT® BPN	-	100%	99.5%	-	-	-	-
95% Confidence Interval	BACT/ALERT® BPA	-	98.2%-100%	97.2%-99.9%	-	-	-	-
	BACT/ALERT® BPN	-	98.2%-100%	97.2%-99.9%	-	-	-	-

\* Ten replicates each of BACT/ALERT® BPA and BACT/ALERT® BPN bottles were inoculated with each microorganism at the inoculum level indicated.

† One BACT/ALERT® BPA and one BACT/ALERT® BPN bottle for each system was inoculated with 10 mL of “unseeded” apheresis platelets to serve as negative controls each time seeded testing was performed. These negative controls served as a sterility test of the platelet units and to establish that LRAP at a 10 mL inoculum volume do not cause false positive results to occur. The number of negative controls tested was supplemented from leftover platelets from the seeded testing such that 208 repetitions of negative controls (104 BACT/ALERT® BPA and 104 BACT/ALERT® BPN) were performed on BACT/ALERT® 3D and 211 repetitions of negative controls (106 BACT/ALERT® BPA and 105 BACT/ALERT® BPN) were performed on BACT/ALERT® VIRTUO®. For BACT/ALERT® 3D, 207 of 208 negative controls were negative (sterile), i.e., there was 1 false positive. For BACT/ALERT® VIRTUO®, all (211) negative controls were negative (sterile), i.e. there were no false positives.

‡ Mean of 20 replicates

§ Mean of 19 replicates

¶ Strict Anaerobe - not tested in aerobic culture bottles (BACT/ALERT® BPA).

\*\* Strict Aerobe - not tested in anaerobic culture bottles (BACT/ALERT® BPN).

The seeded bottle data were analyzed using a non-inferiority approach with a -5% non-inferiority margin based on the overall detection rate.

For BACT/ALERT® BPA, BACT/ALERT® VIRTUO® detected 199/200 (99.5%, CI 97.2% - 99.9%) and BACT/ALERT® 3D detected 200/200 (100%, CI 98.2% - 100%) of the seeded bottles. For the one BACT/ALERT® VIRTUO® status negative bottle, there was no microorganism seen on Gram stain and no colonies on subculture. Repeat testing duplicates were status positive and were considered in agreement with the expected result.

The difference of the two detection rates (BACT/ALERT® VIRTUO® – BACT/ALERT® 3D) is -0.5% with a 95% confidence interval for the difference of the two detection rates of [(-3.2%) - 1.9%], using the Newcombe hybrid score method with continuity correction. Since the lower bound of the confidence interval is > -5%, the BACT/ALERT® VIRTUO® detection rate is considered acceptable.

For BACT/ALERT® BPN, BACT/ALERT® VIRTUO® detected 199/200 (99.5%, CI 97.2% - 99.9%) and BACT/ALERT® 3D detected 200/200 (100%, CI 98.2% - 100%) of the seeded bottles. For the one BACT/ALERT® VIRTUO® status negative bottle, there was no microorganism seen on Gram stain and no colonies on subculture. Repeat testing duplicates were status positive and were considered in agreement with the expected result.

The difference of the two detection rates (BACT/ALERT® VIRTUO® – BACT/ALERT® 3D) is -0.5% with a 95% confidence interval for the difference of the two detection rates of [(-3.2%) - 1.9%], using the Newcombe hybrid score method with continuity correction. Since the lower bound of the confidence interval is > -5%, the BACT/ALERT® VIRTUO® detection rate is considered acceptable.

**BACT/ALERT® 3D and BACT/ALERT® VIRTUO® Systems Comparative Data**

An internal study was conducted for analytical sensitivity during instrument validation to establish substantial equivalence between the BACT/ALERT® VIRTUO® and the BACT/ALERT® 3D Microbial Detection Systems. The LRAP in plasma units were obtained from the inventory of a platelet collection agency. Aliquots of the platelets were seeded with BIOBALL®

microorganisms, and from stock cultures, at a target inoculum of 5 CFU/mL or with microorganisms prepared from serial dilutions at a target inoculum of 10 CFU/mL. The aliquots from platelets were seeded with 13 individual microorganisms:

- *Bacillus cereus* NCTC 7464 Product 56023 BIOBALL® SINGLESHOT
- *Clostridium perfringens* NCTC 8798 Product 56028 BIOBALL® SINGLESHOT
- *Enterobacter cloacae* ATCC® 35549™
- *Escherichia coli* NCTC 12241 Product 56035 BIOBALL® SINGLESHOT
- *Pseudomonas aeruginosa* NCTC 12924 Product 56040 BIOBALL® SINGLESHOT
- *Klebsiella pneumoniae* ATCC® 35657™
- *Proteus mirabilis* ATCC® 7002™
- *Salmonella enterica* subsp. *enterica* serovar Pomona ATCC® 10729™
- *Salmonella enterica* subsp. *enterica* serotype Typhimurium NCTC 12023 Product 56044 BIOBALL® SINGLESHOT
- *Serratia marcescens* ATCC® 43862™
- *Staphylococcus aureus* NCTC 10788 Product 56045 BIOBALL® SINGLESHOT
- *Staphylococcus epidermidis* NCTC 6513 Product 56092 BIOBALL® SINGLESHOT
- *Streptococcus sanguinis* ATCC® 10556™

Four (4) mL of the seeded LRAP aliquots were inoculated into the appropriate BACT/ALERT® BPA and BPN bottles and the bottles were tested on 3 BACT/ALERT® VIRTUO® and 1 BACT/ALERT® 3D. A total of 2 repetitions/system were performed for each microorganism/bottle type with LRAP and system (3 BACT/ALERT® VIRTUO® and 1 BACT/ALERT® 3D). Recovery and time to detection for test microorganisms by bottle type and instrument are presented in the following table, as well as the actual inoculum ranges.

**Table 13: Analytical Sensitivity: Growth Performance in BACT/ALERT® BPA and BACT/ALERT® BPN Culture Bottles**

Microorganism	BACT/ALERT® Culture Bottle	BACT/ALERT® 3D				BACT/ALERT® VIRTUO®			
		% Recovery (n=2)	Mean Inoculum (CFU/mL)*	Time to Detection (hours)		% Recovery (n=6)	Mean Inoculum (CFU/mL)*	Time to Detection (hours)	
				Mean	Range			Mean	Range
<i>Bacillus cereus</i> NCTC 7464	BACT/ALERT® BPA	100.0	5	10.4	10.3-10.6	100.0	5	7.5	7.2-7.9
	BACT/ALERT® BPN	100.0	5	13.8	13.7-13.9	100.0	5	11.5	9.5-12.7
<i>Clostridium perfringens</i> † NCTC 8798	BACT/ALERT® BPA	-	-	-	-	-	-	-	-
	BACT/ALERT® BPN	100.0	1	10.2	10.1-10.3	100.0	1	7.5	7.0-7.8
<i>Enterobacter cloacae</i> ATCC® 35549™	BACT/ALERT® BPA	100.0	1	22.7	22.1-23.3	100.0	1	22.5	18.3-26.6
	BACT/ALERT® BPN	100.0	1	13.2	13.2‡	100.0	1	10.0	9.8-10.1
<i>Escherichia coli</i> NCTC 12241	BACT/ALERT® BPA	100.0	1	13.0	12.7-13.2	100.0	1	9.8	9.7-10.0
	BACT/ALERT® BPN	100.0	1	11.6	11.5-11.8	100.0	1	9.0	8.9-9.2
<i>Klebsiella pneumoniae</i> ATCC® 35657™	BACT/ALERT® BPA	100.0	1	12.0	12.0‡	100.0	1	9.4	9.1-9.7
	BACT/ALERT® BPN	100.0	1	12.1	12.0-12.2	100.0	1	9.6	9.4-9.7
<i>Proteus mirabilis</i> ATCC® 7002™	BACT/ALERT® BPA	100.0	3	13.1	13.0-13.2	100.0	3	10.8	10.3-11.4
	BACT/ALERT® BPN	100.0	3	11.6	11.5-11.8	100.0	3	9.2	8.9-9.6



Microorganism	BACT/ALERT® Culture Bottle	BACT/ALERT® 3D				BACT/ALERT® VIRTUO®			
		% Recovery (n=2)	Mean Inoculum (CFU/mL)*	Time to Detection (hours)		% Recovery (n=6)	Mean Inoculum (CFU/mL)*	Time to Detection (hours)	
				Mean	Range			Mean	Range
<i>Pseudomonas aeruginosa</i> § NCTC 12924	BACT/ALERT® BPA	100.0	1	17.0	16.6-17.5	100.0	1	13.8	13.2 - 14.6
	BACT/ALERT® BPN	-	-	-	-	-	-	-	-
<i>Salmonella enterica</i> subsp. <i>enterica</i> serovar Pomona ATCC® 10729™	BACT/ALERT® BPA	100.0	4	12.8	12.7-13.0	100.0	4	10.1	9.9-10.3
	BACT/ALERT® BPN	100.0	4	11.6	11.5-11.8	100.0	4	9.3	9.1-9.4
<i>Salmonella enterica</i> subsp. <i>enterica</i> serotype Typhimurium NCTC 12023	BACT/ALERT® BPA	100.0	1	13.9	13.9‡	100.0	1	10.8	10.5-11.3
	BACT/ALERT® BPN	100.0	1	12.1	12.0-12.2	100.0	1	9.6	9.0-10.1
<i>Serratia marcescens</i> ATCC® 43862™	BACT/ALERT® BPA	100.0	3	12.1	12.0-12.2	100.0	3	9.7	9.6-9.9
	BACT/ALERT® BPN	100.0	3	12.5	12.5‡	100.0	3	10.1	9.8-10.3
<i>Staphylococcus aureus</i> NCTC 10788	BACT/ALERT® BPA	100.0	1	16.8	16.6-17.0	100.0	1	14.1	13.3-15.1
	BACT/ALERT® BPN	100.0	1	18.1	17.8-18.5	100.0	1	14.3	13.9-14.9
<i>Staphylococcus epidermidis</i> NCTC 6513	BACT/ALERT® BPA	100.0	1	18.7	18.5-19.0	100.0	1	14.9	14.2-15.3
	BACT/ALERT® BPN	100.0	1	22.2	21.8-22.6	100.0	1	18.0	17.4-18.7
<i>Streptococcus sanguinis</i> ATCC® 10556™	BACT/ALERT® BPA	100.0	7	19.3	19.2-19.4	100.0	7	15.6	15.3-16.3
	BACT/ALERT® BPN	100.0	7	21.8	21.6 - 22.1	100.0	7	19.0	16.9-21.6

\* Four negative controls were tested at a volume of 10 mL of unseeded LRAP per bottle type per bottle lot randomly across the four instruments (3 BACT/ALERT® VIRTUO® and 1 BACT/ALERT® 3D) to serve as negative controls each time seeded testing was performed. These negative controls also served as a sterility test of the platelet units. In addition, 4 bottles of each bottle type per each lot were loaded to establish that LRAP at a 10 mL inoculum volume do not cause false positive results to occur. A total of 10 BACT/ALERT® BPA and 10 BACT/ALERT® BPN bottles were tested on the BACT/ALERT® VIRTUO® and 2 BACT/ALERT® BPA and 2 BACT/ALERT® BPN bottles were tested on the BACT/ALERT® 3D. All negative controls were declared negative by both the BACT/ALERT® VIRTUO® and BACT/ALERT® 3D and were negative upon subculture, i.e. there were no false positives.

† Strict Anaerobe - not tested in aerobic culture bottles (BACT/ALERT® BPA).

‡ Both replicates have the same time-to-detection value.

§ Strict Aerobe - not tested in anaerobic culture bottles (BACT/ALERT® BPN).

#### Within-Laboratory Precision (Repeatability) of BACT/ALERT® BPA and BACT/ALERT® BPN Culture Bottles

An internal study was conducted during instrument validation to establish evidence of repeatability of growth performance of the BACT/ALERT® VIRTUO® when tested with 4 mL of LRAP over 10 days with BACT/ALERT® BPA and BACT/ALERT® BPN culture bottles. Bottles were also tested on the BACT/ALERT® 3D, for reference. The platelets were obtained from the

inventory of a platelet collection agency. Aliquots of the platelets were seeded with BIOBALL® microorganisms at a target inoculum of 5 CFU/mL. The aliquots from platelets were seeded with 6 individual microorganisms:

- *Bacillus cereus* NCTC 7464 Product 56023 BIOBALL® SINGLESHOT
- *Escherichia coli* NCTC 12241 Product 56035 BIOBALL® SINGLESHOT
- *Pseudomonas aeruginosa* NCTC 12924 Product 56040 BIOBALL® SINGLESHOT
- *Clostridium perfringens* NCTC 8798 Product 56028 BIOBALL® SINGLESHOT
- *Staphylococcus aureus* NCTC 10788 Product 56045 BIOBALL® SINGLESHOT
- *Streptococcus pyogenes* NCTC 12696 Product 56046 BIOBALL® SINGLESHOT

**Table 14: Within-Laboratory Precision (Repeatability) of BACT/ALERT® BPA and BACT/ALERT® BPN Culture Bottles**

Microorganism	BACT/ALERT® Culture Bottle	BACT/ALERT® 3D				BACT/ALERT® VIRTUO®			
		% Recovery (n=20)*	Mean Inoculum (CFU/mL)†	Time to Detection (hours)		% Recovery (n=60)*	Mean Inoculum (CFU/mL)†	Time to Detection (hours)	
				Mean	Range			Mean	Range
<i>Bacillus cereus</i> NCTC 7464	BACT/ALERT® BPA	100.0	3	10.0	9.6-10.8	100.0	3	7.5	7.0-8.2
	BACT/ALERT® BPN	100.0	3	14.7	11.3-16.3	100.0	3	11.1	8.1-13.4
<i>Clostridium perfringens</i> ‡ NCTC 8798	BACT/ALERT® BPA	-	-	-	-	-	-	-	-
	BACT/ALERT® BPN	95.0	3	16.3	9.6-102.2§	100.0	3	11.4	6.8-62.4§
<i>Escherichia coli</i> NCTC 12241	BACT/ALERT® BPA	100.0	3	12.8	12.2-13.7	100.0	3	9.8	9.2-11.2
	BACT/ALERT® BPN	100.0	3	11.7	11.3-12.2	100.0	3	9.1	8.6-9.7
<i>Pseudomonas aeruginosa</i> ¶ NCTC 12924	BACT/ALERT® BPA	100.0	< 1	17.3	16.8-18.2	98.3	< 1	13.9	13.1-15.4
	BACT/ALERT® BPN	-	-	-	-	-	-	-	-
<i>Staphylococcus aureus</i> NCTC 10788	BACT/ALERT® BPA	100.0	3	16.5	15.6-18.0	100.0	3	13.8	12.4-15.1
	BACT/ALERT® BPN	100.0	3	17.6	16.8-18.2	100.0	3	14.2	13.1-15.4
<i>Streptococcus pyogenes</i> NCTC 12696	BACT/ALERT® BPA	100.0	3	16.3	15.6-17.0	100.0	3	13.6	12.2-15.5
	BACT/ALERT® BPN	100.0	3	12.9	12.5-13.7	100.0	3	10.3	9.6-11.0

\* The number of replicates for *P. aeruginosa* and *C. perfringens* were 20 for the BACT/ALERT® 3D and 60 for the BACT/ALERT® VIRTUO®.

† A total of 250 negative control bottles were tested with volumes from 4-10 mL of unseeded LRAP (188 on BACT/ALERT® VIRTUO® (96 BACT/ALERT® BPA and 92 BACT/ALERT® BPN) and 62 on BACT/ALERT® 3D (31 BACT/ALERT® BPA and 31 BACT/ALERT® BPN)) to assess the risk of false positives. All negative controls were declared negative by the instruments after 7 days and upon subculture. No false positive bottles were observed.

‡ Strict Anaerobe - not tested in aerobic culture bottles (BACT/ALERT® BPA).

§ *C. perfringens* demonstrated sporadic, prolonged time-to-detection in some bottles on each detection system. Bottles were subcultured and identified as *C. perfringens*.

¶ Strict Aerobe - not tested in anaerobic culture bottles (BACT/ALERT® BPN).

## Limited Warranty

bioMérieux warrants the performance of the product for its stated intended use provided that all procedures for usage, storage and handling, shelf life (when applicable), and precautions are strictly followed as detailed in the instructions for use (IFU).













Except as expressly set forth above, bioMérieux hereby disclaims all warranties, including any implied warranties of merchantability and fitness for a particular purpose or use, and disclaims all liability, whether direct, indirect or consequential, for any use of the reagent, software, instrument and disposables (the "System") other than as set forth in the IFU.


## Availability

bioMérieux BACT/ALERT® BPA 100/case	REF 279044
---	------------

For technical assistance in the USA, contact bioMérieux Customer Service at 1-800-634-7656. Outside the USA, contact your local bioMérieux Representative.

## Index of Symbols

Symbol	Meaning
	Catalogue number
	Manufacturer
	Date of manufacture
	Temperature limit
	Use by date
	Batch code
	Consult Instructions for Use
	Contains sufficient for <n> tests
	Authorized Representative in the European Community
	This way up
	<i>In Vitro</i> Diagnostic Medical Device
	Do not reuse

Symbol	Meaning
	Latex-free

Instructions for use provided in the kit or downloadable from <http://www.biomerieux.com>.

## Revision History

### Change type categories

N/A	Not applicable (First publication)
Correction	Correction of documentation anomalies
Technical change	Addition, revision and/or removal of information related to the product
Administrative	Implementation of non-technical changes noticeable to the user

**Note:** *Minor typographical, grammar, and formatting changes are not included in the revision history.*

Release Date	Part Number	Change Type	Change Summary
2022-03	050794-04	Technical	<b>Specimen Collection and Preparation</b> - added Note: "Some literature suggests..."
2020-10	050794-03	Technical	<b>Intended Use</b> <ul style="list-style-type: none"> <li>Inserted new IU statement for harmonization with other BPA docs</li> </ul>
			<b>Performance Characteristics</b> <ul style="list-style-type: none"> <li><i>Propionibacterium</i> sp. changed to <i>Cutibacterium</i> sp.</li> <li>Added section "Equivalency of Platelet Preparation Method..." and Table 7.</li> </ul>

Release Date	Part Number	Change Type	Change Summary
2019-04	9315715-C	Correction	<b>Performance Characteristics -</b> <ul style="list-style-type: none"> <li>• Correction of names for following microorganisms:                             <ul style="list-style-type: none"> <li>◦ <i>Salmonella enterica</i> subsp. <i>enterica</i> serovar Heidelberg (Tables 5, 6, 9, and 10)</li> <li>◦ <i>Salmonella enterica</i> subsp. <i>enterica</i> serovar Pomona (Table 12)</li> <li>◦ <i>Salmonella enterica</i> subsp. <i>enterica</i> serotype Typhimurium (Table 12)</li> </ul> </li> <li>• Correction of ATCC numbers for following microorganisms:                             <ul style="list-style-type: none"> <li>◦ <i>Enterobacter cloacae</i> (Table 12)</li> <li>◦ <i>Klebsiella pneumoniae</i> (Table 12)</li> </ul> </li> <li>• Correction of percent recovery for <i>Clostridium perfringens</i> (Table 13)</li> </ul>
		Technical change	<b>Reagents -</b> Addition of alcohol pads or equivalent to list of additional materials required <b>BACT/ALERT® BPA Culture Bottle Test Procedure -</b> Addition of precaution regarding personal protective equipment (PPE) <b>Specimen Test/Inoculation Procedure -</b> Addition of step regarding proper mixing after inoculation
		Administrative	<b>Limited Warranty -</b> Addition of statement
		2017-09	9306999-B
		Administrative	<b>Index of Symbols -</b> Revision to reflect new symbols on product

BIOMERIEUX, the BIOMERIEUX logo, BACT/ALERT, BIOBALL, and VIRTUO are used, pending, and/or registered trademarks belonging to bioMérieux, or one of its subsidiaries, or one of its companies.

This product may be protected by one or more patents, see: <http://www.biomerieux-usa.com/patents>.

The ATCC trademark and trade name and any and all ATCC catalog numbers are trademarks of the American Type Culture Collection.

CLSI is a registered trademark of Clinical and Laboratory Standards Institute.

Any other name or trademark is the property of its respective owner.

©BIOMÉRIEUX 2005, 2008, 2010, 2017, 2019, 2020, 2022

