



BẢN SAO

IN CYTO

Quality Assurance Sheet

Inspector	Manager
	

Product Model : DHC-N01
Lot Number : 6201A233
Date of Quality Test : 01 Apr. 2016

Parts	Object	Quality Evaluation							Remarks
		Valid limits	Results					Ave.	
Chamber	Chamber height(A)	95~105 microns	99	98	99	97	99	98	OK
	Chamber height(B)	95~105 microns	96	97	96	98	97	97	
Contact angle		Less than 60°	55.8	52.5	54.1	59.1	54.0	55.1	OK
Grid	Shape error rate	Less than 5 %	0.20%						OK
Final product	Output error rate	Less than 5 %	0.50%						OK

We certify;

that this quality assurance sheet is issued on a basis of our precise QA Process

using appropriate measurement instruments and methods.

CHỨNG THỰC BẢN SAO ĐÚNG VỚI BẢN CHÍNH

NGÀY: 25-08-2017

SỐ CHỨNG THỰC.....QUYỀN SỐ.....SCT/BS

2669 04

Manager : Young Dook Byeon
Quality Assurance Department

INCYTO Co., Ltd.



INCYTO Co., Ltd. Chonan Plant

112, Seonggeo-gil, Songgo-eup, Seobuk-gu, Cheonan-si, Chungnam-do, 31044, Korea

TEL +82-41-550-9761 FAX +82-41-622-9985

CÔNG CHỨNG VIÊN
Chu Cảnh Hưng