

FUJIFILM



SAO Y BẢN CHÍNH
CERTIFIED TRUE COPY
FUJIFILM Corporation
26-30, NISHIAZABU 2-CHOME, MINATO-KU
TOKYO 106-8620, JAPAN

Declaration of Conformity

05 - CEI - 02 - E

Manufacturer: FUJIFILM Corporation
Address: 26-30, Nishiazabu 2-chome, Minatoku,
TOKYO 106-8620, JAPAN
European Representative: FUJIFILM Europe GmbH
Address: HEESENSTRASSE 31,
40549 DÜSSELDORF, GERMANY
Product: **FUJI DRI-CHEM SLIDE Na-K-CI**
GMDN Code 30196,30199,30200
EDMA Code 11.04.01.03,11.04.01.06,11.04.01.07
Start of CE Marking: Lot No.106804
Classification: The device is not referred in Article 9 (2) and (3).
Conformity Assessment Route: Annex III without Section 6

We herewith declare that the above mentioned product meets the provisions of the following EC Council Directive.

Directive:

COUNCIL DIRECTIVE 98/79/EC of 27 October 1998 on in vitro diagnostic medical devices (IVDD 98/79/EC)

Standards:

- EN ISO13485:2012 Medical devices - Quality management systems - Requirements for regulatory purpose
- EN ISO 14971:2012 Medical devices - Application of risk management to medical devices
- EN 13612: 2002 Performance evaluation of in vitro diagnostic medical devices
- EN 13640:2002 Stability testing of in vitro diagnostic reagents
- EN ISO 18113-1:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 1: Terms, definitions and general requirements
- EN ISO 18113-2:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use
- EN ISO 15223-1:2012 Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied -- Part 1: General requirements

Place: JAPAN

Date: 2014-10-01

A handwritten signature in black ink, appearing to read "Naotake Mitsumori".

Naotake Mitsumori
Senior Manager,
Quality Assurance and Regulatory Affairs Division
Medical Systems Business Division
FUJIFILM Corporation