



SAO Y BẢN CHÍNH  
CERTIFIED TRUE COPY

FUJIFILM Corporation  
26-30, NISHIAZABU 2-CHOME, MINATO-KU  
TOKYO 106-8620, JAPAN

## Declaration of Conformity

05 - CEI - 355 - F

Manufacturer: FUJIFILM Corporation  
Address: 26-30, Nishiazabu 2-chome, Minatoku,  
TOKYO 106-8620, JAPAN  
European Representative: FUJIFILM Europe GmbH  
Address: HEESENSTRASSE 31,  
40549 DÜSSELDORF, GERMANY  
Product: **FUJI DRI-CHEM SLIDE ALP-P III**  
GMDN Code 33165  
EDMA Code 11.01.01.05  
Start of CE Marking: Lot No.121507  
Classification: The device is not referred in Article 9 (2) and (3).  
Conformity Assessment Route: Annex III without Section 6

We herewith declare that the above mentioned product meets the provisions of the following EC Council Directive.

Directive:


COUNCIL DIRECTIVE 98/79/EC of 27 October 1998 on in vitro diagnostic medical devices (IVDD 98/79/EC)

Standards:

- EN ISO13485:2012 Medical devices - Quality management systems - Requirements for regulatory purpose
- EN ISO 14971:2012 Medical devices - Application of risk management to medical devices
- EN 13612: 2002 Performance evaluation of in vitro diagnostic medical devices
- EN 13640:2002 Stability testing of in vitro diagnostic reagents
- EN ISO 18113-1:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 1: Terms, definitions and general requirements
- EN ISO 18113-2:2011 In vitro diagnostic medical devices - Information supplied by the manufacturer (labelling) - Part 2: In vitro diagnostic reagents for professional use
- EN ISO 15223-1:2012 Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied -- Part 1: General requirements

Place: JAPAN

Date: 2014-10-01

  
Naotake Mitsumori  
Senior Manager,  
Quality Assurance and Regulatory Affairs Division  
Medical Systems Business Division  
FUJIFILM Corporation