

Certificate of Compliance



No. 0D151214.JTMQQ85
 Technical Construction File no. LC-15001240

Certificate's Holder: Xuzhou Lianchuang Medical Equipment Co., Ltd.
 1#F1,2#F2, No.40, Minyuan Road, Tongshan District, Xuzhou, Jiangsu, China

Certification ECM Mark:



Product: Colposcope
Model(s): LC-9100A, LC-9100B, LC-9100C, LC-9200A, LC-9200B, LC-9200C

Verification to: Standard:
 EN 60601-1:2006+A1:2013,
 EN 60601-2-18:2015,
 EN ISO 10993-1:2009+AC:2010,
 EN 1041:2008, EN ISO 15223-1:2012

related to CE Directive(s):
 93/42/EEC (Medical Devices)

Remark: The product(s) has been verified on a voluntary basis. The product(s) satisfies the requirements of the Certification Mark of ECM, in reference to the above listed Standard(s). The above Certification Mark can be affixed on the product(s) accordingly to the ECM regulation about its release and its use. Regulation can be found at www.entecerma.it.

Whereas the Manufacturer is responsible of the CE certification of the product(s) and not exempted to perform all the necessary activities before placing the product(s) on the market.

The Manufacturer is also responsible to maintain efficient the internal production control to ensure the product(s) are in compliance with the Certification ECM Mark.

This certificate can be checked for validity at www.entecerma.it

Date of issue 14 December 2015

Expiry date 13 December 2020

Chief Manager
 Tim Mahan



Deputy Manager
 Viola Miller



Ente Certificazione Macchine Srl

Via Ca' Bella, 243 - Loc. Castello di Serravalle - 40053 Valsamoggia (BO) - ITALY
 ☎ +39 051 6705141 📠 +39 051 6705156 ✉ info@entecerma.it 🌐 www.entecerma.it